



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: JIGNESHDate: 03/24/2017Run Number: LB86524BalanceID: WC SC-3DATE & TIME IN: 03/24/2017 09:00DATE & TIME OUT: 03/24/2017 13:15OVEN TEMP: 104 °COVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB86524BL	LB86524BL	1.4582	1.4582	100	1.4582	1.4582	1.4582	0	0
2	LB86524BS	LB86524BS	1.3901	1.3901	100	1.4451	1.4451	1.4451	0.055	550
3	I2373-01	SPDES-1	1.4055	1.4055	400	1.4058	1.4058	1.4058	0.0003	0.7
4	I2373-01DUP	SPDES-1DUP	1.4308	1.4308	400	1.4312	1.4312	1.4312	0.0004	1

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

1386524

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS I2373

WorkList ID : 96792

Date : 3/24/2017 8:25:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2373-01-D	TSS	Cool 4 deg C	SKAN06	M21	SPDES-1	03/23/2017	SM2540 D

Date/Time 03/24/17 8:25 AM
 Received by: JP
 Relinquished by: CD

Date/Time 03/24/17 10:00 AM
 Received by: OR
 Relinquished by: JO