



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB86529

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	03/24/2017	08:30
2	I2351-02		1	>100.0	>212.0	03/24/2017	09:00
3	I2351-02DUP		1	>100.0	>212.0	03/24/2017	09:20

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 86529

WorkList Name : FP32417

WorkList ID : 96799

Date : 3/24/2017 08:00 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/24/2017	Water	I2351-02	Flash Point	Cool 4 deg C	PSEG01	K42	MANHOLE-001	03/21/2017	1010A

Date/Time 3/24/17 8:05 AM

Received by: KS

Relinquished by: CP

Date/Time 3/24/17 9:30 AM

Received by: CP

Relinquished by: KS