

7470A


LB 86532

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35973



A= 0.0000e+000

B= 2.7788e-004

C= -2.5074e-002

Rho= 0.9999680

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	12	0.000	12				
0.2	0.200	0.177	-0.023	727	0.0 %	727				
2.5	2.500	2.527	0.027	9185	0.0 %	9185				
5.0	5.000	5.051	0.051	18268	0.0 %	18268				
7.5	7.500	7.495	-0.005	27064	0.0 %	27064				
10.0	10.000	9.971	-0.029	35973	0.0 %	35973				

LB86532 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Mar 2017 12:56:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7470A	PPB	24 Mar 2017 12:58:20	S	Std
0.2 - 1	50.2	727	-	0.2000	7470A	PPB	24 Mar 2017 13:00:28	S	Std
2.5 - 1	52.5	9185	-	2.5000	7470A	PPB	24 Mar 2017 13:02:37	S	Std
5.0 - 1	55	18268	-	5.0000	7470A	PPB	24 Mar 2017 13:04:45	S	Std
7.5 - 1	57.5	27064	-	7.5000	7470A	PPB	24 Mar 2017 13:06:53	S	Std
10.0 - 1	50	35973	-	10.0000	7470A	PPB	24 Mar 2017 13:12:08	S	Std
ICV86 - 1	ICV86	13693	3.7799	-	7470A	PPB	24 Mar 2017 13:14:42	U	SMPL
ICB86 - 1	ICB86	-56	-0.0406	-	7470A	PPB	24 Mar 2017 13:16:48	U	SMPL
CCV62 - 1	CCV62	18352	5.0745	-	7470A	PPB	24 Mar 2017 13:18:54	U	SMPL
CCB62 - 1	CCB62	-55	-0.0404	-	7470A	PPB	24 Mar 2017 13:21:01	U	SMPL
CRA - 1	CRA	703	0.1703	-	7470A	PPB	24 Mar 2017 13:23:07	U	SMPL
HighStd - 1	HighStd	35885	9.9465	-	7470A	PPB	24 Mar 2017 13:25:14	U	SMPL
ChkStd - 1	ChkStd	24703	6.8393	-	7470A	PPB	24 Mar 2017 13:27:21	U	SMPL
PB97781BL - 1	PBW	-70	-0.0445	-	7470A	PPB	24 Mar 2017 13:29:29	U	SMPL
PB97781BS - 1	LCSW	13486	3.7224	-	7470A	PPB	24 Mar 2017 13:31:36	U	SMPL
PB97781TB - 1	PB97747TB	-50	-0.0390	-	7470A	PPB	24 Mar 2017 13:33:44	U	SMPL
I2351-02 - 1	MANHOLE-001	46	-0.0123	-	7470A	PPB	24 Mar 2017 13:35:52	U	SMPL
I2351-02DUP - 1	MANHOLE-001DUP	-1	-0.0254	-	7470A	PPB	24 Mar 2017 13:38:00	U	SMPL
I2351-02MS - 1	MANHOLE-001MS	14029	3.8733	-	7470A	PPB	24 Mar 2017 13:40:09	U	SMPL
I2351-02MSD - 1	MANHOLE-001MSD	13706	3.7835	-	7470A	PPB	24 Mar 2017 13:42:16	U	SMPL
CCV63 - 1	CCV63	18066	4.9951	-	7470A	PPB	24 Mar 2017 13:44:23	U	SMPL
CCB63 - 1	CCB63	-62	-0.0423	-	7470A	PPB	24 Mar 2017 13:46:29	U	SMPL
I2364-04 - 1	WC-COMP-10	0	-0.0251	-	7470A	PPB	24 Mar 2017 13:48:36	U	SMPL
I2364-05 - 1	WC-COMP-11	12	-0.0217	-	7470A	PPB	24 Mar 2017 13:50:43	U	SMPL
I2364-06 - 1	WC-COMP-12	5	-0.0237	-	7470A	PPB	24 Mar 2017 13:52:49	U	SMPL
PB97782BL - 1	PBW	12	-0.0217	-	7470A	PPB	24 Mar 2017 13:54:56	U	SMPL
PB97782BS - 1	LCSW	13880	3.8319	-	7470A	PPB	24 Mar 2017 13:57:04	U	SMPL
I2367-14 - 1	E2-GWTS-3-21	69	-0.0059	-	7470A	PPB	24 Mar 2017 13:59:11	U	SMPL
I2371-01 - 1	S8-0668	88	-0.0006	-	7470A	PPB	24 Mar 2017 14:05:53	U	SMPL
I2371-01DUP - 1	S8-0668DUP	68	-0.0062	-	7470A	PPB	24 Mar 2017 14:08:00	U	SMPL
I2371-01MS - 1	S8-0668MS	13681	3.7766	-	7470A	PPB	24 Mar 2017 14:10:09	U	SMPL
I2371-01MSD - 1	S8-0668MSD	13715	3.7860	-	7470A	PPB	24 Mar 2017 14:12:16	U	SMPL
CCV64 - 1	CCV64	18228	5.0401	-	7470A	PPB	24 Mar 2017 14:14:23	U	SMPL
CCB64 - 1	CCB64	-37	-0.0354	-	7470A	PPB	24 Mar 2017 14:16:30	U	SMPL
I2371-02 - 1	S8-0665	93	0.0008	-	7470A	PPB	24 Mar 2017 14:18:37	U	SMPL
I2371-03 - 1	S8-0669	74	-0.0045	-	7470A	PPB	24 Mar 2017 14:20:44	U	SMPL
I2351-02LX5 - 1		11	-0.0220	-	7470A	PPB	24 Mar 2017 14:22:51	U	SMPL
I2351-02A - 1		14892	4.1131	-	7470A	PPB	24 Mar 2017 14:24:57	U	SMPL
I2371-01LX5 - 1		57	-0.0092	-	7470A	PPB	24 Mar 2017 14:27:05	U	SMPL
I2371-01A - 1		14643	4.0439	-	7470A	PPB	24 Mar 2017 14:29:13	U	SMPL
CCV65 - 1	CCV65	18374	5.0806	-	7470A	PPB	24 Mar 2017 14:31:21	U	SMPL
CCB65 - 1	CCB65	-32	-0.0340	-	7470A	PPB	24 Mar 2017 14:33:29	U	SMPL