


ISM02.4

LB 86539
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47343



A= 0.0000e+000

B= 2.1147e-004

C= -1.8594e-002

Rho= 0.9999902

Accept= Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.005	-0.005	66	0.000	66					-3
0.2	0.200	0.195	-0.005	1011	0.0 %	1011					1
2.5	2.500	2.523	0.023	12020	0.0 %	12020					1
5.0	5.000	4.975	-0.025	23614	0.0 %	23614					0
7.5	7.500	7.518	0.018	35642	0.0 %	35642					0
10.0	10.000	9.993	-0.007	47343	0.0 %	47343					0

noB
3/24/17

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	66	-		0 ISM02.4	PPB	3/24/2017 15:38	Std
	0.2 S0.2	1011	-		0.2 ISM02.4	PPB	3/24/2017 15:40	Std
	2.5 S2.5	12020	-		2.5 ISM02.4	PPB	3/24/2017 15:44	Std
	5 S5	23614	-		5 ISM02.4	PPB	3/24/2017 15:46	Std
	7.5 S7.5	35642	-		7.5 ISM02.4	PPB	3/24/2017 15:48	Std
	10 S10	47343	-		10 ISM02.4	PPB	3/24/2017 15:51	Std
ICV	ICV	18571	3.9085 -		ISM02.4	PPB	3/24/2017 15:53	SMPL
ICB	ICB	-7	-0.0201 -		ISM02.4	PPB	3/24/2017 15:56	SMPL
CCV066	CCV066	23805	5.0153 -		ISM02.4	PPB	3/24/2017 15:58	SMPL
CCB066	CCB066	-9	-0.0205 -		ISM02.4	PPB	3/24/2017 16:00	SMPL
PB97785BL	PBS004	34	-0.0114 -		ISM02.4	PPB	3/24/2017 16:02	SMPL
I2344-01	MDACH1	1395	0.2764 -		ISM02.4	PPB	3/24/2017 16:05	SMPL
I2344-02	MDACH1D	1169	0.2286 -		ISM02.4	PPB	3/24/2017 16:07	SMPL
I2344-03	MDACH1S	13009	2.7324 -		ISM02.4	PPB	3/24/2017 16:09	SMPL
I2344-04	MDACH4	890	0.1696 -		ISM02.4	PPB	3/24/2017 16:12	SMPL
I2344-05	MDACH5	948	0.1819 -		ISM02.4	PPB	3/24/2017 16:14	SMPL
I2344-06	MDACH6	584	0.1049 -		ISM02.4	PPB	3/24/2017 16:16	SMPL
I2344-07	MDACH7	653	0.1195 -		ISM02.4	PPB	3/24/2017 16:18	SMPL
I2344-08	MDACJ0	950	0.1823 -		ISM02.4	PPB	3/24/2017 16:21	SMPL
I2344-09	MDACJ1	426	0.0715 -		ISM02.4	PPB	3/24/2017 16:23	SMPL
I2344-10	MDACH2	907	0.1732 -		ISM02.4	PPB	3/24/2017 16:25	SMPL
I2344-11	MDACH3	952	0.1827 -		ISM02.4	PPB	3/24/2017 16:28	SMPL
I2344-12	MDACH8	1384	0.2741 -		ISM02.4	PPB	3/24/2017 16:30	SMPL
I2344-13	MDACH9	1905	0.3842 -		ISM02.4	PPB	3/24/2017 16:32	SMPL
I2344-14	MDACJ4	1526	0.3041 -		ISM02.4	PPB	3/24/2017 16:34	SMPL
I2344-15	MDACG7	9236	1.9345 -		ISM02.4	PPB	3/24/2017 16:37	SMPL
I2344-16	MDACG8	465613	HIGH -		ISM02.4	PPB	3/24/2017 16:43	
CCV067	CCV067	22490	4.7373 -		ISM02.4	PPB	3/24/2017 16:49	SMPL
CCB067	CCB067	-53	-0.0298 -		ISM02.4	PPB	3/24/2017 16:51	SMPL
I2344-16DLX10	MDACG8	70636	14.9185 -		ISM02.4	PPB	3/24/2017 16:56	SMPL
CCV068	CCV068	23607	4.9735 -		ISM02.4	PPB	3/24/2017 17:01	SMPL
CCB068	CCB068	-19	-0.0226 -		ISM02.4	PPB	3/24/2017 17:03	SMPL
I2344-16DLX50	MDACG8	14186	2.9813 -		ISM02.4	PPB	3/24/2017 17:05	SMPL
CCV069	CCV069	23644	4.9813 -		ISM02.4	PPB	3/24/2017 17:12	SMPL
CCB069	CCB069	7	-0.0171 -		ISM02.4	PPB	3/24/2017 17:14	SMPL