



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86544

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 99.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/24/2017	16:45
2	CAL2	1	Water			7.02	03/24/2017	16:46
3	CAL3	1	Water			10.03	03/24/2017	16:47
4	ICV	1	Water			7.01	03/24/2017	16:48
5	CCV1	1	Water			2.01	03/24/2017	16:50
6	I2384-03	1	Solid	20.02	20	8.03	03/24/2017	16:51
7	I2384-03DUP	1	Solid	20.04	20	8.04	03/24/2017	16:52
8	I2384-04	1	Solid	20.05	20	8.26	03/24/2017	16:53
9	CCV2	1	Water			12.02	03/24/2017	16:54

2386544

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2384 WorkList ID : 96820 Date : 3/24/2017 4:31:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2384-03	Corrosivity	Cool 4 deg C	ENTA04		WC-B42-B51-001	03/23/2017	9045C
	Solid	I2384-04	Corrosivity	Cool 4 deg C	ENTA04	M41	WC-B43-B50-001	03/23/2017	9045C

Date/Time 03/24/17 11:31 AM
Received by: JA
Relinquished by: J.C.

Date/Time 03/24/17 5:00 PM
Received by: J.C.
Relinquished by: J.S.