

245.1


LB 86 545

INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

34233



A= 0.0000e+000

B= 2.9073e-004

C= 1.0742e-002

Rho= 0.9998274

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.013 | 0.013  | 8     | 0.000      | 8     |       |       |       |       |
| 0.2    | 0.200  | 0.207 | 0.007  | 676   | 0.0 %      | 676   |       |       |       |       |
| 2.5    | 2.500  | 2.514 | 0.014  | 8610  | 0.0 %      | 8610  |       |       |       |       |
| 5.0    | 5.000  | 4.886 | -0.114 | 16770 | 0.0 %      | 16770 |       |       |       |       |
| 7.5    | 7.500  | 7.616 | 0.116  | 26160 | 0.0 %      | 26160 |       |       |       |       |
| 10.0   | 10.000 | 9.963 | -0.037 | 34233 | 0.0 %      | 34233 |       |       |       |       |

## LB86545 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Mar 2017 19:22:04

| Sample ID       | Extended ID            | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|------------------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                     | 8      | -       | 0.0000   | 245.1  | PPB   | 24 Mar 2017 19:23:59 | S    | Std  |
| 0.2 - 1         | 50.2                   | 676    | -       | 0.2000   | 245.1  | PPB   | 24 Mar 2017 19:26:07 | S    | Std  |
| 2.5 - 1         | 52.5                   | 8610   | -       | 2.5000   | 245.1  | PPB   | 24 Mar 2017 19:28:15 | S    | Std  |
| 5.0 - 1         | 55                     | 16770  | -       | 5.0000   | 245.1  | PPB   | 24 Mar 2017 19:30:23 | S    | Std  |
| 7.5 - 1         | 57.5                   | 26160  | -       | 7.5000   | 245.1  | PPB   | 24 Mar 2017 19:32:31 | S    | Std  |
| 10.0 - 1        | 510                    | 34233  | -       | 10.0000  | 245.1  | PPB   | 24 Mar 2017 19:34:38 | S    | Std  |
| ICV87 - 1       | ICV87                  | 13240  | 3.8600  | -245.1   |        | PPB   | 24 Mar 2017 19:37:19 | U    | SMPL |
| ICB87 - 1       | ICB87                  | -45    | -0.0023 | -245.1   |        | PPB   | 24 Mar 2017 19:39:25 | U    | SMPL |
| CCV70 - 1       | CCV70                  | 16760  | 4.8834  | -245.1   |        | PPB   | 24 Mar 2017 19:41:32 | U    | SMPL |
| CCB70 - 1       | CCB70                  | -59    | -0.0064 | -245.1   |        | PPB   | 24 Mar 2017 19:43:38 | U    | SMPL |
| CRA - 1         | CRA                    | 658    | 0.2020  | -245.1   |        | PPB   | 24 Mar 2017 19:45:44 | U    | SMPL |
| HighStd - 1     | HighStd                | 34555  | 10.0569 | -245.1   |        | PPB   | 24 Mar 2017 19:47:51 | U    | SMPL |
| ChkStd - 1      | ChkStd                 | 23382  | 6.8086  | -245.1   |        | PPB   | 24 Mar 2017 19:49:58 | U    | SMPL |
| PB97770BL - 1   | PBW                    | -77    | -0.0116 | -245.1   |        | PPB   | 24 Mar 2017 19:52:06 | U    | SMPL |
| PB97770BS - 1   | LCSW                   | 13504  | 3.9367  | -245.1   |        | PPB   | 24 Mar 2017 19:54:13 | U    | SMPL |
| I2273-01 - 1    | EFFLUENT-DAY1METALS    | 109    | 0.0424  | -245.1   |        | PPB   | 24 Mar 2017 20:00:53 | U    | SMPL |
| I2273-01DUP - 1 | EFFLUENT-DAY1METALSDUP | 79     | 0.0337  | -245.1   |        | PPB   | 24 Mar 2017 20:03:01 | U    | SMPL |
| I2273-01MS - 1  | EFFLUENT-DAY1METALSMS  | 11741  | 3.4242  | -245.1   |        | PPB   | 24 Mar 2017 20:05:09 | U    | SMPL |
| I2273-01MSD - 1 | EFFLUENT-DAY1METALSMSD | 12605  | 3.6754  | -245.1   |        | PPB   | 24 Mar 2017 20:07:18 | U    | SMPL |
| I2355-01 - 1    | 032117-PRRT-E-D-M      | 9      | 0.0134  | -245.1   |        | PPB   | 24 Mar 2017 20:09:25 | U    | SMPL |
| CCV71 - 1       | CCV71                  | 15819  | 4.6098  | -245.1   |        | PPB   | 24 Mar 2017 20:11:32 | U    | SMPL |
| CCB71 - 1       | CCB71                  | -109   | -0.0209 | -245.1   |        | PPB   | 24 Mar 2017 20:13:38 | U    | SMPL |
| I2355-02 - 1    | 032117-PRRT-E-U-M      | 33     | 0.0203  | -245.1   |        | PPB   | 24 Mar 2017 20:15:46 | U    | SMPL |
| I2356-01 - 1    | 032017-PRRT-E-U-M      | 31     | 0.0198  | -245.1   |        | PPB   | 24 Mar 2017 20:17:53 | U    | SMPL |
| I2356-02 - 1    | 032017-PRRT-E-D-M      | 26     | 0.0183  | -245.1   |        | PPB   | 24 Mar 2017 20:20:00 | U    | SMPL |
| I2273-01LX5 - 1 |                        | 78     | 0.0334  | -245.1   |        | PPB   | 24 Mar 2017 20:22:06 | U    | SMPL |
| I2273-01A - 1   |                        | 12418  | 3.6210  | -245.1   |        | PPB   | 24 Mar 2017 20:24:14 | U    | SMPL |
| CCV72 - 1       | CCV72                  | 16101  | 4.6918  | -245.1   |        | PPB   | 24 Mar 2017 20:26:22 | U    | SMPL |
| CCB72 - 1       | CCB72                  | -109   | -0.0209 | -245.1   |        | PPB   | 24 Mar 2017 20:28:30 | U    | SMPL |