



Analytical Summary Report

Analysis Method: 120.1

ANALYST: AMANDEEP

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB86549

Reagent/Standard	Lot/Log #
conductivity standard	WP49968

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB86549BL		1	24.5	0.91	0.92	03/27/2017	10:59
2	LB86549BS	1413	1	22.5	1302.00	1367.29	03/27/2017	11:02
3	I2212-08		1	22.4	699.00	735.53	03/27/2017	11:05

Result = $\frac{\text{(Reading)}}{1 + 0.0191 \text{ (Temperature - 25)}}$