



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB86551

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP53788
Residual chlorine std, 0PPM	WP53784
Residual chlorine std, 0.1PPM	WP53785
Residual chlorine std, 0.2PPM	WP53786
Residual chlorine std, and CCV , 0.4PPM	WP53787
Residual chlorine std, 1.6PPM	WP53789
Residual chlorine ICV-LCS, 0.4PPM	WP53790
Residual chlorine std, Intermediate 10PPM	WP53782
Residual chlorine std, Intermediate-SS 10PPM	WP53783

Intercept: 0.0143

Slope: 0.9401

Regression: 0.999424

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	03/27/2017	11:35
2	CAL2	0.1	1	0.000	0.100	0.100	03/27/2017	11:40
3	CAL3	0.2	1	0.000	0.200	0.200	03/27/2017	11:45
4	CAL4	0.4	1	0.000	0.400	0.400	03/27/2017	11:50
5	CAL5	0.8	1	0.000	0.800	0.800	03/27/2017	11:55
6	CAL6	1.6	1	0.000	1.500	1.500	03/27/2017	12:00



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB86551

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.400	0.400	0.400	03/27/2017	12:05
2	ICB				1	0.000	0.000	0.000	0.000	03/27/2017	12:10
3	CCV1			0.4	1	0.000	0.400	0.400	0.400	03/27/2017	12:15
4	CCB1				1	0.000	0.000	0.000	0.000	03/27/2017	12:20
5	LB86551BL	50	50		1	0.000	0.000	0.000	0.000	03/27/2017	12:25
6	LB86551BS	50	50	0.4	1	0.000	0.390	0.390	0.390	03/27/2017	12:30
7	I2212-19	50	50		1	0.000	1.980	1.980	1.980	03/27/2017	12:34
8	I2212-19				2	0.000	1.010	1.010	1.010	03/27/2017	12:38
9	CCV2			0.4	1	0.000	0.410	0.410	0.410	03/27/2017	12:42
10	CCB2				1	0.000	0.000	0.000	0.000	03/27/2017	12:47