


245.1

LB 86553
INSTRUMENT INDEX

Linear

 μ Abs.:

32962



A= 0.0000e+000

B= 3.0265e-004

C= 1.4070e-002

Rho= 0.9999799

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	9	0.000	9				
0.2	0.200	0.204	0.004	628	0.0 %	628				
2.5	2.500	2.479	-0.021	8146	0.0 %	8146				
5.0	5.000	4.970	-0.030	16375	0.0 %	16375				
7.5	7.500	7.540	0.040	24865	0.0 %	24865				
10.0	10.000	9.990	-0.010	32962	0.0 %	32962				

LB86553 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Mar 2017 12:15:50

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	9	-	0.0000245.1	PPB	27 Mar 2017 12:22:03	S	Std
0.2 - 1	50.2	628	-	0.2000245.1	PPB	27 Mar 2017 12:24:11	S	Std
2.5 - 1	53.5	8146	-	2.5000245.1	PPB	27 Mar 2017 12:26:20	S	Std
5.0 - 1	55	16375	-	5.0000245.1	PPB	27 Mar 2017 12:28:28	S	Std
7.5 - 1	57.5	24865	-	7.5000245.1	PPB	27 Mar 2017 12:30:35	S	Std
10.0 - 1	511	32962	-	10.0000245.1	PPB	27 Mar 2017 12:32:44	S	Std
ICV89 - 1	ICV89	12579	3.8211	-245.1	PPB	27 Mar 2017 12:35:22	U	SMPL
ICB89 - 1	ICB89	-81	-0.0104	-245.1	PPB	27 Mar 2017 12:37:28	U	SMPL
CCV77 - 1	CCV77	16123	4.8937	-245.1	PPB	27 Mar 2017 12:39:36	U	SMPL
CCB77 - 1	CCB77	-82	-0.0107	-245.1	PPB	27 Mar 2017 12:41:42	U	SMPL
CRA - 1	CRA	618	0.2011	-245.1	PPB	27 Mar 2017 12:43:49	U	SMPL
HighStd - 1	HighStd	33239	10.0739	-245.1	PPB	27 Mar 2017 12:45:56	U	SMPL
ChkStd - 1	ChkStd	23418	7.1016	-245.1	PPB	27 Mar 2017 12:48:03	U	SMPL
PB97813BL - 1	PBW	-102	-0.0168	-245.1	PPB	27 Mar 2017 12:50:12	U	SMPL
PB97813BS - 1	LCSW	14146	4.2954	-245.1	PPB	27 Mar 2017 12:52:21	U	SMPL
I2387-01 - 1	032317-PRRT-E-D-M	5	0.0156	-245.1	PPB	27 Mar 2017 12:54:29	U	SMPL
I2387-02 - 1	032317-PRRT-E-U-M	136	0.0552	-245.1	PPB	27 Mar 2017 12:56:37	U	SMPL
I2388-01 - 1	032217-PRRT-E-D-M	31	0.0235	-245.1	PPB	27 Mar 2017 12:58:45	U	SMPL
I2388-02 - 1	032217-PRRT-E-U-M	185	0.0701	-245.1	PPB	27 Mar 2017 13:00:54	U	SMPL
I2388-02DUP - 1	032217-PRRT-E-U-MDUP	136	0.0552	-245.1	PPB	27 Mar 2017 13:03:01	U	SMPL
CCV78 - 1	CCV78	16083	4.8816	-245.1	PPB	27 Mar 2017 13:05:08	U	SMPL
CCB78 - 1	CCB78	-48	-0.0005	-245.1	PPB	27 Mar 2017 13:07:15	U	SMPL
I2388-02MS - 1	032217-PRRT-E-U-MMS	16430	4.9866	-245.1	PPB	27 Mar 2017 13:09:22	U	SMPL
I2388-02MSD - 1	032217-PRRT-E-U-MMSD	15824	4.8032	-245.1	PPB	27 Mar 2017 13:11:29	U	SMPL
I2388-02LX5 - 1		-17	0.0089	-245.1	PPB	27 Mar 2017 13:13:35	U	SMPL
I2388-02A - 1		13954	4.2373	-245.1	PPB	27 Mar 2017 13:15:43	U	SMPL
CCV79 - 1	CCV79	16199	4.9167	-245.1	PPB	27 Mar 2017 13:17:50	U	SMPL
CCB79 - 1	CCB79	-67	-0.0062	-245.1	PPB	27 Mar 2017 13:19:58	U	SMPL