



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : _____

Run Number: LB86562

Temp.: 22.1 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/24/2017	18:35
2	CAL2	1	Water			7.02	03/24/2017	18:36
3	CAL3	1	Water			10.03	03/24/2017	18:37
4	ICV	1	Water			7.00	03/24/2017	18:38
5	CCV1	1	Water			2.02	03/24/2017	18:39
6	I2421-01	1	Water			6.69	03/24/2017	18:40
7	I2421-01DUP	1	Water			6.70	03/24/2017	18:41
8	I2421-03	1	Water			6.68	03/24/2017	18:42
9	CCV2	1	Water			12.02	03/24/2017	18:43

1386562

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 032417

WorkList ID : 96823

Date : 3/24/2017 6:30:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	12421-01-G	pH	Cool 4 deg C	DEWB01	M41	HHD-GW1	03/24/2017	SM 4500-PH B
	Water	12421-03-G	pH	Cool 4 deg C	DEWB01	M41	HHD-GW2	03/24/2017	SM 4500-PH B

Date/Time 03/24/17 6:30 PM
Received by: JP
Relinquished by:

Date/Time 03/24/17 7:50 PM
Received by: JP
Relinquished by: JP