



## Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86565

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.8

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2183         |
| buffer solution pH 7 yellow    | W2159         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2154         |
| BUFFER PH 7.00 GREEN 1PINT PK6 | W2140         |
| pH 2.00 Buffer                 | W2079         |
| Buffer Solution, PH12 (500ml)  | W2099         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  |             |             | 4.01        | 03/27/2017 | 13:35     |
| 2   | CAL2        | 1  | Water  |             |             | 7.00        | 03/27/2017 | 13:36     |
| 3   | CAL3        | 1  | Water  |             |             | 10.01       | 03/27/2017 | 13:37     |
| 4   | ICV         | 1  | Water  |             |             | 7.00        | 03/27/2017 | 13:38     |
| 5   | CCV1        | 1  | Water  |             |             | 2.02        | 03/27/2017 | 13:39     |
| 6   | I2422-02    | 1  | Solid  | 20.02       | 20          | 8.20        | 03/27/2017 | 13:40     |
| 7   | I2422-02DUP | 1  | Solid  | 20.04       | 20          | 8.21        | 03/27/2017 | 13:41     |
| 8   | CCV2        | 1  | Water  |             |             | 12.02       | 03/27/2017 | 13:42     |

# WORKLIST(Hardcopy Internal Chain)

1386565

WorkList Name : CORROSIVITY I2422      WorkList ID : 96866      Date : 3/27/2017 1:23:59 PM

| Due Date | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|----------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
|          | Solid  | I2422-02 | Corrosivity | Cool 4 deg C | DEWB01   | M41              | HHD-S1          | 03/23/2017   | 9045C  |

Date/Time 03/27/17 1:23PM  
 Received by: JS  
 Relinquished by: CP

Date/Time 03/27/17 5:30PM  
 Received by: CP  
 Relinquished by: JS