


7471A

LB86568
INSTRUMENT ID: CV2

Linear

 μ Abs.:

44479



A= 0.0000e+000

B= 2.2406e-004

C= 5.0640e-002

Rho= 0.9996850

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.062	0.062	49	0.000	49				
0.2	0.200	0.218	0.018	746	0.0 %	746				
2.5	2.500	2.496	-0.004	10916	0.0 %	10916				
5.0	5.000	4.807	-0.193	21230	0.0 %	21230				
7.5	7.500	7.600	0.100	33693	0.0 %	33693				
10.0	10.000	10.017	0.017	44479	0.0 %	44479				

LB86568 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 27 Mar 2017 14:05:03

Sample ID	Extended ID	μ Abs.	Conc.	Std. Cond.	Method	Units	Date	Type	Type
0.0 - 1	50	49	-	0.0000	7471A	PPB	27 Mar 2017 14:07:43	S	Std
0.2 - 1	50.2	746	-	0.2000	7471A	PPB	27 Mar 2017 14:10:00	S	Std
2.5 - 1	52.5	10916	-	2.5000	7471A	PPB	27 Mar 2017 14:12:18	S	Std
5.0 - 1	55	21230	-	5.0000	7471A	PPB	27 Mar 2017 14:14:35	S	Std
7.5 - 1	57.5	33693	-	7.5000	7471A	PPB	27 Mar 2017 14:19:35	S	Std
10.0 - 1	510	44479	-	10.0000	7471A	PPB	27 Mar 2017 14:21:52	S	Std
ICV90 - 1	ICV90	16398	3.7248	-	7471A	PPB	27 Mar 2017 14:24:44	U	SMPL
ICB90 - 1	ICB90	-52	0.0390	-	7471A	PPB	27 Mar 2017 14:26:59	U	SMPL
CCV80 - 1	CCV80	21539	4.8767	-	7471A	PPB	27 Mar 2017 14:29:14	U	SMPL
CCB80 - 1	CCB80	-63	0.0365	-	7471A	PPB	27 Mar 2017 14:31:30	U	SMPL
CRA - 1	CRA	726	0.2133	-	7471A	PPB	27 Mar 2017 14:33:46	U	SMPL
HighStd - 1	HighStd	44006	9.9107	-	7471A	PPB	27 Mar 2017 14:36:02	U	SMPL
ChkStd - 1	ChkStd	29888	6.7474	-	7471A	PPB	27 Mar 2017 14:38:18	U	SMPL
PB97817BL - 1	PBS	5	0.0518	-	7471A	PPB	27 Mar 2017 14:40:35	U	SMPL
PB97817BS - 1	LCSS	16587	3.7672	-	7471A	PPB	27 Mar 2017 14:42:53	U	SMPL
I2381-01 - 1	3-4-DENSEGRADEAGGREGATE-RCA	3641	0.8664	-	7471A	PPB	27 Mar 2017 14:45:10	U	SMPL
I2381-02 - 1	1-1-2-DENSEGRADEAGGREGATE-RCA	1077	0.2920	-	7471A	PPB	27 Mar 2017 14:47:27	U	SMPL
I2394-01 - 1	COMP1	985	0.2713	-	7471A	PPB	27 Mar 2017 14:49:45	U	SMPL
I2394-02 - 1	COMP2	663	0.1992	-	7471A	PPB	27 Mar 2017 14:52:03	U	SMPL
I2394-03 - 1	COMP3	284	0.1143	-	7471A	PPB	27 Mar 2017 14:54:19	U	SMPL
CCV81 - 1	CCV81	21882	4.9536	-	7471A	PPB	27 Mar 2017 14:56:35	U	SMPL
CCB81 - 1	CCB81	-50	0.0394	-	7471A	PPB	27 Mar 2017 14:58:51	U	SMPL
I2394-04 - 1	COMP4	645	0.1952	-	7471A	PPB	27 Mar 2017 15:01:07	U	SMPL
I2394-05 - 1	COMP5	541	0.1719	-	7471A	PPB	27 Mar 2017 15:03:23	U	SMPL
I2394-06 - 1	COMP6	441	0.1495	-	7471A	PPB	27 Mar 2017 15:05:39	U	SMPL
I2413-01 - 1	TS-01	3334	0.7977	-	7471A	PPB	27 Mar 2017 15:07:55	U	SMPL
I2413-02 - 1	TS-02	868	0.2451	-	7471A	PPB	27 Mar 2017 15:10:11	U	SMPL
I2414-01 - 1	E2-S13-BASE	10840	2.4795	-	7471A	PPB	27 Mar 2017 15:12:28	U	SMPL
I2414-01DUP - 1	E2-S13-BASEDUP	11325	2.5881	-	7471A	PPB	27 Mar 2017 15:14:46	U	SMPL
I2414-01MS - 1	E2-S13-BASEMS	27343	6.1772	-	7471A	PPB	27 Mar 2017 15:17:03	U	SMPL
I2414-01MSD - 1	E2-S13-BASEMSD	27465	6.2045	-	7471A	PPB	27 Mar 2017 15:19:21	U	SMPL
I2414-02 - 1	E2-R13-BASE	31191	7.0393	-	7471A	PPB	27 Mar 2017 15:21:37	U	SMPL
CCV82 - 1	CCV82	21787	4.9323	-	7471A	PPB	27 Mar 2017 15:23:54	U	SMPL
CCB82 - 1	CCB82	-25	0.0450	-	7471A	PPB	27 Mar 2017 15:26:10	U	SMPL
I2414-03 - 1	E2-S12-BASE	8609	1.9796	-	7471A	PPB	27 Mar 2017 15:28:26	U	SMPL
I2414-04 - 1	E2-P13-WSW	9774	2.2406	-	7471A	PPB	27 Mar 2017 15:30:43	U	SMPL
I2414-01LX5 - 1		2211	0.5460	-	7471A	PPB	27 Mar 2017 15:32:59	U	SMPL
I2414-01A - 1		29013	6.5513	-	7471A	PPB	27 Mar 2017 15:35:16	U	SMPL
CCV83 - 1	CCV83	21863	4.9493	-	7471A	PPB	27 Mar 2017 15:37:32	U	SMPL
CCB83 - 1	CCB83	-8	0.0488	-	7471A	PPB	27 Mar 2017 15:39:49	U	SMPL