



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86569

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP53792
COD CCV std, 50ppm	WP53791

Temp In (C): <u>150</u>	Date In: <u>03/27/2017</u>	Time In: <u>13:45</u>
Temp Out (C): <u>150</u>	Date Out: <u>03/27/2017</u>	Time Out: <u>15:45</u>

Intercept: -0.5466

Slope: 1.0088

Regression: 1

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	03/16/2017	13:50
2	CAL2	10	1	Water	9.000	03/16/2017	13:50
3	CAL3	50	1	Water	50.000	03/16/2017	13:51
4	CAL4	100	1	Water	100.000	03/16/2017	13:51
5	CAL5	150	1	Water	151.000	03/16/2017	13:52



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86569

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	51.000	51.097	03/16/2017	13:52
2	ICB				1	Water	0.000	0.542	03/16/2017	13:53
3	CCV1	50			1	Water	53.000	53.080	03/27/2017	16:40
4	CCB1				1	Water	0.000	0.542	03/27/2017	16:40
5	LB86569BL				1	Water	0.000	0.542	03/27/2017	16:40
6	LB86569BS	50			1	Water	49.000	49.114	03/27/2017	16:41
7	I2361-02				1	Water	0.000	0.542	03/27/2017	16:41
8	I2361-02DUP				1	Water	0.000	0.542	03/27/2017	16:42
9	I2361-02MS	50			1	Water	51.000	51.097	03/27/2017	16:42
10	I2361-02MSD	50			1	Water	51.000	51.097	03/27/2017	16:43
11	I2361-03				1	Water	1.000	1.533	03/27/2017	16:44
12	I2421-01				1	Water	51.000	51.097	03/27/2017	16:44
13	I2421-03				1	Water	48.000	48.123	03/27/2017	16:45
14	CCV2	50			1	Water	52.000	52.088	03/27/2017	16:45
15	CCB2				1	Water	0.000	0.542	03/27/2017	16:46

WORKLIST(Hardcopy Internal Chain)

2B86569

WorkList Name : COD032717

WorkList ID : 96859

Date : 3/27/2017 11:18:16 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2361-02	COD	Conc H2SO4 tc	SHAW22	M21	OUTFALL001-170314	03/14/2017	SM5220 D
	Water	I2361-03	COD	Conc H2SO4 tc	SHAW22	M21	OUTFALL001-170321	03/21/2017	SM5220 D
	Water	I2421-01	COD	Conc H2SO4 tc	DEWB01	M41	HHD-GW1	03/24/2017	SM5220 D
	Water	I2421-03	COD	Conc H2SO4 tc	DEWB01	M41	HHD-GW2	03/24/2017	SM5220 D

Date/Time 03/27/17 13:21
 Received by: IB
 Relinquished by: CR

Date/Time 03/27/17 14:58
 Received by: CP
 Relinquished by: IB