



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 03/28/2017Run Number: LB86574DATE & TIME IN: 03/27/2017 17:25OVEN TEMP: 98 °CBalanceID: WC SC-3DATE & TIME OUT: 03/28/2017 12:55OVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @105°C (g)	2nd Dish+Sample weight after Drying @105°C (g)	Final Dish+Sample weight after Drying @105°C (g)	Weight (g)	Result (mg/L)
1	LB86574BL	LB86574BL	85.0136	85.0136	100	85.0136	85.0136	85.0136	0	0
2	I2433-01	NORTHPORTALWALL-HYDRODEMO	75.2538	75.2538	100	75.566	75.566	75.566	0.3122	3122
3	I2433-01DUP	NORTHPORTALWALL-HYDRODEMODU	68.9641	68.9641	100	69.2767	69.2767	69.2767	0.3126	3126

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \star 1000 \star 1000$

2086574

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS I2433

WorkList ID : 96878

Date : 3/27/2017 5:03:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2433-01	TS	Cool 4 deg C	TULL01	M51	NORTHPORTALWALL-HYDR	03/27/2017	SM2540 B

Date/Time 03/27/17 15:00 PM
Received by: SP
Relinquished by: LC

Date/Time 03/27/17 6:30 PM
Received by: LC
Relinquished by: SP