


7471A

LB86578
INSTRUMENT ID: CV4

Linear

 μ Abs.:

43187



A= 0.0000e+000

B= 2.3289e-004

C= 1.7097e-002

Rho= 0.9999047

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	33	0.000	33				
0.2	0.200	0.230	0.030	913	0.0 %	913				
2.5	2.500	2.465	-0.035	10510	0.0 %	10510				
5.0	5.000	4.990	-0.010	21355	0.0 %	21355				
7.5	7.500	7.415	-0.085	31768	0.0 %	31768				
10.0	10.000	10.075	0.075	43187	0.0 %	43187				

LB86578 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 28 Mar 2017 09:37:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	33	-	0.0000	7471A	PPB	28 Mar 2017 09:50:52	S	Std
0.2 - 1	S0.2	913	-	0.2000	7471A	PPB	28 Mar 2017 09:53:09	S	Std
2.5 - 1	S2.5	10510	-	2.5000	7471A	PPB	28 Mar 2017 09:55:27	S	Std
5.0 - 1	S5	21355	-	5.0000	7471A	PPB	28 Mar 2017 09:57:44	S	Std
7.5 - 1	S7.5	31768	-	7.5000	7471A	PPB	28 Mar 2017 10:00:02	S	Std
10.0 - 1	S10	43187	-	10.0000	7471A	PPB	28 Mar 2017 10:04:54	S	Std
ICV91 - 1	ICV91	17120	4.0041	-7471A	PPB	PPB	28 Mar 2017 10:07:42	U	SMPL
ICB91 - 1	ICB91	-42	0.0073	-7471A	PPB	PPB	28 Mar 2017 10:09:57	U	SMPL
CCV84 - 1	CCV84	21756	5.0838	-7471A	PPB	PPB	28 Mar 2017 10:12:13	U	SMPL
CCB84 - 1	CCB84	-32	0.0096	-7471A	PPB	PPB	28 Mar 2017 10:14:28	U	SMPL
CRA - 1	CRA	890	0.2244	-7471A	PPB	PPB	28 Mar 2017 10:16:44	U	SMPL
HighStd - 1	HighStd	43297	10.1005	-7471A	PPB	PPB	28 Mar 2017 10:19:00	U	SMPL
ChkStd - 1	ChkStd	31228	7.2897	-7471A	PPB	PPB	28 Mar 2017 10:21:16	U	SMPL
PB97857BL - 1	PBS	-102	-0.0067	-7471A	PPB	PPB	28 Mar 2017 10:23:33	U	SMPL
PB97857BS - 1	LCSS	15149	3.5451	-7471A	PPB	PPB	28 Mar 2017 10:25:50	U	SMPL
I2422-01 - 1	HHD-S1	5417	1.2787	-7471A	PPB	PPB	28 Mar 2017 10:28:07	U	SMPL
I2431-01 - 1	E2-T11-BASE	7670	1.8033	-7471A	PPB	PPB	28 Mar 2017 10:30:24	U	SMPL
I2431-02 - 1	E2-U11-SSW	33736	7.8738	-7471A	PPB	PPB	28 Mar 2017 10:32:42	U	SMPL
I2431-03 - 1	E2-S13-SSW	15947	3.7310	-7471A	PPB	PPB	28 Mar 2017 10:35:00	U	SMPL
I2431-04 - 1	E2-T12-SSW	1584	0.3860	-7471A	PPB	PPB	28 Mar 2017 10:37:16	U	SMPL
CCV85 - 1	CCV85	21671	5.0640	-7471A	PPB	PPB	28 Mar 2017 10:39:32	U	SMPL
CCB85 - 1	CCB85	-19	0.0127	-7471A	PPB	PPB	28 Mar 2017 10:41:49	U	SMPL
I2431-05 - 1	E2-T11-SSW	49869	11.6310	-7471A	PPB	PPB	28 Mar 2017 10:44:05	U	SMPL
I2431-06 - 1	E2-R-S-12-13-BASE	12631	2.9587	-7471A	PPB	PPB	28 Mar 2017 10:46:21	U	SMPL
I2431-06DUP - 1	E2-R-S-12-13-BASEDUP	13201	3.0915	-7471A	PPB	PPB	28 Mar 2017 10:48:37	U	SMPL
I2431-06MS - 1	E2-R-S-12-13-BASEMS	29470	6.8803	-7471A	PPB	PPB	28 Mar 2017 10:50:53	U	SMPL
I2431-06MSD - 1	E2-R-S-12-13-BASEMSD	31296	7.3056	-7471A	PPB	PPB	28 Mar 2017 10:53:10	U	SMPL
I2431-06LX5 - 1		2350	0.5644	-7471A	PPB	PPB	28 Mar 2017 10:55:27	U	SMPL
I2431-06A - 1		30327	7.0799	-7471A	PPB	PPB	28 Mar 2017 10:57:44	U	SMPL
I2431-05DLX5 - 1		9791	2.2973	-7471A	PPB	PPB	28 Mar 2017 11:00:02	U	SMPL
CCV86 - 1	CCV86	21722	5.0759	-7471A	PPB	PPB	28 Mar 2017 11:02:20	U	SMPL
CCB86 - 1	CCB86	-14	0.0138	-7471A	PPB	PPB	28 Mar 2017 11:04:36	U	SMPL