


7470A

LB86582
INSTRUMENT ID: CV4

Linear

 μ Abs.:

36696



A= 0.0000e+000

B= 2.7482e-004

C= 1.7510e-002

Rho= 0.9998240

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	16	0.000	16				
0.2	0.200	0.212	0.012	709	0.0 %	709				
2.5	2.500	2.520	0.020	9106	0.0 %	9106				
5.0	5.000	4.972	-0.028	18028	0.0 %	18028				
7.5	7.500	7.372	-0.128	26760	0.0 %	26760				
10.0	10.000	10.102	0.102	36696	0.0 %	36696				

LB86582 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Mar 2017 11:19:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	16	-	0.0000	7470A	PPB	28 Mar 2017 11:22:20	S	Std
0.2 - 1	S0.2	709	-	0.2000	7470A	PPB	28 Mar 2017 11:24:28	S	Std
2.5 - 1	S2.5	9106	-	2.5000	7470A	PPB	28 Mar 2017 11:26:37	S	Std
5.0 - 1	S5	18028	-	5.0000	7470A	PPB	28 Mar 2017 11:28:45	S	Std
7.5 - 1	S7.5	26760	-	7.5000	7470A	PPB	28 Mar 2017 11:30:53	S	Std
10.0 - 1	S10	36696	-	10.0000	7470A	PPB	28 Mar 2017 11:33:00	S	Std
ICV92 - 1	ICV92	13732	3.7913	-7470A	PPB	PPB	28 Mar 2017 11:35:41	U	SMPL
ICB92 - 1	ICB92	-78	-0.0039	-7470A	PPB	PPB	28 Mar 2017 11:37:47	U	SMPL
CCV87 - 1	CCV87	17883	4.9321	-7470A	PPB	PPB	28 Mar 2017 11:39:54	U	SMPL
CCB87 - 1	CCB87	-85	-0.0058	-7470A	PPB	PPB	28 Mar 2017 11:42:00	U	SMPL
CRA - 1	CRA	680	0.2044	-7470A	PPB	PPB	28 Mar 2017 11:44:07	U	SMPL
HighStd - 1	HighStd	35828	9.8637	-7470A	PPB	PPB	28 Mar 2017 11:46:13	U	SMPL
ChkStd - 1	ChkStd	23847	6.5711	-7470A	PPB	PPB	28 Mar 2017 11:48:20	U	SMPL
PB97852BL - 1	PBW	-111	-0.0130	-7470A	PPB	PPB	28 Mar 2017 11:50:28	U	SMPL
PB97852BS - 1	LCSW	14359	3.9636	-7470A	PPB	PPB	28 Mar 2017 11:52:37	U	SMPL
PB97852TB - 1	PB97815TB	-80	-0.0045	-7470A	PPB	PPB	28 Mar 2017 11:54:44	U	SMPL
I2384-03 - 1	WC-B42-B51-001	511	0.1579	-7470A	PPB	PPB	28 Mar 2017 11:56:53	U	SMPL
I2384-04 - 1	WC-B43-B50-001	805	0.2387	-7470A	PPB	PPB	28 Mar 2017 11:59:00	U	SMPL
I2384-04DUP - 1	WC-B43-B50-001DUP	881	0.2596	-7470A	PPB	PPB	28 Mar 2017 12:01:09	U	SMPL
I2384-04MS - 1	WC-B43-B50-001MS	13459	3.7163	-7470A	PPB	PPB	28 Mar 2017 12:03:16	U	SMPL
CCV88 - 1	CCV88	18119	4.9969	-7470A	PPB	PPB	28 Mar 2017 12:05:23	U	SMPL
CCB88 - 1	CCB88	-108	-0.0122	-7470A	PPB	PPB	28 Mar 2017 12:07:29	U	SMPL
I2384-04MSD - 1	WC-B43-B50-001MSD	14385	3.9708	-7470A	PPB	PPB	28 Mar 2017 12:09:35	U	SMPL
PB97853BL - 1	PBW	-73	-0.0026	-7470A	PPB	PPB	28 Mar 2017 12:11:42	U	SMPL
PB97853BS - 1	LCSW	14381	3.9697	-7470A	PPB	PPB	28 Mar 2017 12:13:49	U	SMPL
I2421-01 - 1	HHD-GW1	462	0.1445	-7470A	PPB	PPB	28 Mar 2017 12:20:35	U	SMPL
I2421-01DUP - 1	HHD-GW1DUP	404	0.1285	-7470A	PPB	PPB	28 Mar 2017 12:22:41	U	SMPL
I2421-01MS - 1	HHD-GW1MS	16119	4.4473	-7470A	PPB	PPB	28 Mar 2017 12:24:49	U	SMPL
I2421-01MSD - 1	HHD-GW1MSD	16319	4.5023	-7470A	PPB	PPB	28 Mar 2017 12:26:57	U	SMPL
I2421-03 - 1	HHD-GW2	2508	0.7068	-7470A	PPB	PPB	28 Mar 2017 12:29:06	U	SMPL
I2421-05 - 1	FB-3-24-17	0	0.0175	-7470A	PPB	PPB	28 Mar 2017 12:31:14	U	SMPL
I2431-07 - 1	E2-FB-3-24	21	0.0233	-7470A	PPB	PPB	28 Mar 2017 12:33:21	U	SMPL
CCV89 - 1	CCV89	17820	4.9148	-7470A	PPB	PPB	28 Mar 2017 12:35:28	U	SMPL
CCB89 - 1	CCB89	-51	0.0035	-7470A	PPB	PPB	28 Mar 2017 12:37:35	U	SMPL
I2384-04LX5 - 1		173	0.0651	-7470A	PPB	PPB	28 Mar 2017 12:39:42	U	SMPL
I2384-04A - 1		15328	4.2299	-7470A	PPB	PPB	28 Mar 2017 12:41:49	U	SMPL
I2421-01LX5 - 1		46	0.0302	-7470A	PPB	PPB	28 Mar 2017 12:49:08	U	SMPL
I2421-01A - 1		16596	4.5784	-7470A	PPB	PPB	28 Mar 2017 12:51:14	U	SMPL
CCV90 - 1	CCV90	17946	4.9494	-7470A	PPB	PPB	28 Mar 2017 12:53:21	U	SMPL
CCB90 - 1	CCB90	-41	0.0062	-7470A	PPB	PPB	28 Mar 2017 12:55:29	U	SMPL