



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86584

Temp.: 21.6 °C

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/28/2017	11:50
2	CAL2	1	Water			7.01	03/28/2017	11:51
3	CAL3	1	Water			10.04	03/28/2017	11:52
4	ICV	1	Water			6.99	03/28/2017	11:53
5	CCV1	1	Water			2.02	03/28/2017	11:54
6	I2434-01	1	Water			4.76	03/28/2017	11:55
7	I2434-01DUP	1	Water			4.77	03/28/2017	11:56
8	CCV2	1	Water			12.01	03/28/2017	11:57

LB86584

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12434

WorkList ID : 96911

Date : 3/28/2017 11:38:32 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	12434-01-A	pH	Cool 4 deg C	STER07	M52	CELL-D	03/27/2017	SM 4500-PH B

Date/Time 03/28/17 11:38 AM
 Received by: JD
 Relinquished by: CD

Date/Time 03/28/17 12:35 PM
 Received by: ed
 Relinquished by: JD