


245.1

LB 86585-
INSTRUMENT ID: CVA

Linear

 μ Abs.:

32925



A= 0.0000e+000

B= 3.0410e-004

C= -2.0056e-002

Rho= 0.9999810

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	23	0.000	23				
0.2	0.200	0.205	0.005	740	0.0 %	740				
2.5	2.500	2.485	-0.015	8238	0.0 %	8238				
5.0	5.000	5.048	0.048	16666	0.0 %	16666				
7.5	7.500	7.482	-0.018	24670	0.0 %	24670				
10.0	10.000	9.993	-0.007	32925	0.0 %	32925				

LB86585 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Mar 2017 13:03:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	245.1	PPB	28 Mar 2017 13:07:00	S	Std
0.2 - 1	50.2	740	-	0.2000	245.1	PPB	28 Mar 2017 13:09:08	S	Std
2.5 - 1	52.5	8238	-	2.5000	245.1	PPB	28 Mar 2017 13:11:17	S	Std
5.0 - 1	55	16666	-	5.0000	245.1	PPB	28 Mar 2017 13:13:25	S	Std
7.5 - 1	57.5	24670	-	7.5000	245.1	PPB	28 Mar 2017 13:15:32	S	Std
10.0 - 1	58	32925	-	10.0000	245.1	PPB	28 Mar 2017 13:17:40	S	Std
ICV93 - 1	ICV93	13226	4.0020	-245.1		PPB	28 Mar 2017 13:20:16	U	SMPL
ICB93 - 1	ICB93	53	-0.0039	-245.1		PPB	28 Mar 2017 13:22:22	U	SMPL
CCV91 - 1	CCV91	16753	5.0746	-245.1		PPB	28 Mar 2017 13:24:29	U	SMPL
CCB91 - 1	CCB91	43	-0.0070	-245.1		PPB	28 Mar 2017 13:26:35	U	SMPL
CRA - 1	CRA	722	0.1995	-245.1		PPB	28 Mar 2017 13:28:43	U	SMPL
HighStd - 1	HighStd	33295	10.1051	-245.1		PPB	28 Mar 2017 13:30:49	U	SMPL
ChkStd - 1	ChkStd	23112	7.0084	-245.1		PPB	28 Mar 2017 13:32:56	U	SMPL
PB97854BL - 1	PBW	-94	-0.0486	-245.1		PPB	28 Mar 2017 13:35:04	U	SMPL
PB97854BS - 1	LCSW	13289	4.0212	-245.1		PPB	28 Mar 2017 13:37:13	U	SMPL
I2387-02 - 1	032317-PRRT-E-U-M	12	-0.0164	-245.1		PPB	28 Mar 2017 13:39:20	U	SMPL
I2388-02 - 1	032217-PRRT-E-U-M	198	0.0402	-245.1		PPB	28 Mar 2017 13:41:29	U	SMPL
I2428-01 - 1	032417-PRRT-F-U-M	45	-0.0064	-245.1		PPB	28 Mar 2017 13:43:37	U	SMPL
I2428-02 - 1	032417-PRRT-F-D-M	48	-0.0055	-245.1		PPB	28 Mar 2017 13:45:45	U	SMPL
I2428-02DUP - 1	032417-PRRT-F-D-MDUP	59	-0.0021	-245.1		PPB	28 Mar 2017 13:47:52	U	SMPL
CCV92 - 1	CCV92	16611	5.0314	-245.1		PPB	28 Mar 2017 13:49:59	U	SMPL
CCB92 - 1	CCB92	57	-0.0027	-245.1		PPB	28 Mar 2017 13:52:06	U	SMPL
I2428-02MS - 1	032417-PRRT-F-D-MMS	15173	4.5941	-245.1		PPB	28 Mar 2017 13:54:13	U	SMPL
I2428-02MSD - 1	032417-PRRT-F-D-MMSD	15661	4.7425	-245.1		PPB	28 Mar 2017 13:56:19	U	SMPL
I2433-01 - 1	NORTHPORTALWALL-HYDRODEMO	1117	0.3196	-245.1		PPB	28 Mar 2017 13:58:26	U	SMPL
I2428-02LX5 - 1		0	-0.0201	-245.1		PPB	28 Mar 2017 14:00:33	U	SMPL
I2428-02A - 1		16319	4.9426	-245.1		PPB	28 Mar 2017 14:02:40	U	SMPL
CCV93 - 1	CCV93	16596	5.0269	-245.1		PPB	28 Mar 2017 14:04:48	U	SMPL
CCB93 - 1	CCB93	63	-0.0009	-245.1		PPB	28 Mar 2017 14:06:56	U	SMPL