

Analysis Method: SM4500-CL C

ANALYST: AMANDEEP

Parameter: Chloride

SUPERVISOR REVIEW BY: apatel

Run Number: LB86586

Constant: 35450

Normality1: 0.0141

Reagent/Standard	Lot/Log #
Chloride LCS std, 100PPM	WP53825
Indicator-acidifier reagent - Chloride	WP53496

Seq	Lab ID	True Value (mg/L)	Dilution	Sample Volume (ml)	Titrant Initial (ml)	Titrant Final (ml)	Titrant Diff. (ml)	Result (mg/L)	Anal Date	Anal Time
1	LB86586BL		1	50	0.00	0.10	0.10	0.00	03/28/2017	10:00
2	LB86586BS	100	1	10	0.44	2.50	2.06	97.97	03/28/2017	10:10
3	I2433-01		1	10	2.50	3.40	0.90	39.99	03/28/2017	10:19
4	I2433-01DUP		1	10	3.40	4.26	0.86	37.99	03/28/2017	10:28
5	I2433-01MS	100	1	10	4.26	7.00	2.74	131.96	03/28/2017	10:37
6	I2433-01MSD	100	1	10	7.00	9.78	2.78	133.96	03/28/2017	10:46

Titrant Diff = TitrantFinal - TitrantInitial

Result = (Titrant Diff - Row Blank Value * Normality)) * Constant /

0886586

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CL032817

WorkList ID : 96900

Date : 3/28/2017 9:27:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2433-01	Chloride	Cool 4 deg C	TULL01	M51	NORTHPORTALWALL-HYDRI	03/27/2017	SM4500-CL C

Date/Time 3-28-17 9:30

Received by: AK

Relinquished by: [Signature]

Date/Time 3-28-17 11:30

Received by: [Signature]

Relinquished by: AK