


ISM02.4

LB 86587
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47810



A= 0.0000e+000

B= 2.0940e-004

C= -9.4211e-003

Rho= 0.9999972

Accept=Accepted

Slope
-1/2105 accept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.00	0.000	-0.001	-0.001	40	0.000	40					-
0.2	0.200	0.194	-0.006	973	0.0 %	973					-3
2.5	2.500	2.503	0.003	12000	0.0 %	12000					0
5.0	5.000	5.015	0.015	23995	0.0 %	23995					0
7.5	7.500	7.486	-0.014	35797	0.0 %	35797					0
10.0	10.000	10.002	0.002	47810	0.0 %	47810					0

NB
3/28/17

LB86587
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	40	-		0 ISM02.4	PPB	3/28/2017 15:32	Std
	0.2 S0.2	973	-		0.2 ISM02.4	PPB	3/28/2017 15:34	Std
	2.5 S2.5	12000	-		2.5 ISM02.4	PPB	3/28/2017 15:36	Std
	5 S5	23995	-		5 ISM02.4	PPB	3/28/2017 15:38	Std
	7.5 S7.5	35797	-		7.5 ISM02.4	PPB	3/28/2017 15:41	Std
	10 S10	47810	-		10 ISM02.4	PPB	3/28/2017 15:43	Std
ICV	ICV	18640	3.8938 -		ISM02.4	PPB	3/28/2017 15:46	SMPL
ICB	ICB	-54	-0.0207 -		ISM02.4	PPB	3/28/2017 15:48	SMPL
CCV094	CCV094	23874	4.9897 -		ISM02.4	PPB	3/28/2017 15:50	SMPL
CCB094	CCB094	-58	-0.0216 -		ISM02.4	PPB	3/28/2017 15:53	SMPL
PB97855BL	PBS004	32	-0.0027 -		ISM02.4	PPB	3/28/2017 15:55	SMPL
I2404-01	MDACC4	1162	0.2339 -		ISM02.4	PPB	3/28/2017 15:57	SMPL
I2404-02	MDACC4D	1075	0.2157 -		ISM02.4	PPB	3/28/2017 16:00	SMPL
I2404-03	MDACC4S	12584	2.6256 -		ISM02.4	PPB	3/28/2017 16:02	SMPL
I2404-04	MDACC5	893	0.1776 -		ISM02.4	PPB	3/28/2017 16:04	SMPL
I2404-05	MDACC6	920	0.1832 -		ISM02.4	PPB	3/28/2017 16:06	SMPL
I2404-06	MDACC7	248	0.0425 -		ISM02.4	PPB	3/28/2017 16:09	SMPL
I2404-07	MDACC8	209	0.0343 -		ISM02.4	PPB	3/28/2017 16:11	SMPL
I2404-08	MDACC9	206	0.0337 -		ISM02.4	PPB	3/28/2017 16:13	SMPL
I2404-09	MDACD0	504	0.0961 -		ISM02.4	PPB	3/28/2017 16:15	SMPL
I2404-10	MDACD1	1268	0.2561 -		ISM02.4	PPB	3/28/2017 16:18	SMPL
I2404-11	MDACD2	2228	0.4571 -		ISM02.4	PPB	3/28/2017 16:20	SMPL
I2404-12	MDACD3	1834	0.3746 -		ISM02.4	PPB	3/28/2017 16:22	SMPL
I2404-13	MDACD4	2501	0.5143 -		ISM02.4	PPB	3/28/2017 16:25	SMPL
I2404-14	MDACE7	602	0.1166 -		ISM02.4	PPB	3/28/2017 16:27	SMPL
I2404-15	MDACE8	1156	0.2326 -		ISM02.4	PPB	3/28/2017 16:29	SMPL
I2404-16	MDACE9	690	0.1351 -		ISM02.4	PPB	3/28/2017 16:31	SMPL
I2404-17	MDACF0	895	0.178 -		ISM02.4	PPB	3/28/2017 16:34	SMPL
I2404-18	MDACCO	10694	2.2299 -		ISM02.4	PPB	3/28/2017 16:36	SMPL
PB97856BL	PBS004	-22	-0.014 -		ISM02.4	PPB	3/28/2017 16:38	SMPL
CCV095	CCV095	23802	4.9747 -		ISM02.4	PPB	3/28/2017 16:41	SMPL
CCB095	CCB095	-66	-0.0232 -		ISM02.4	PPB	3/28/2017 16:43	SMPL
I2407-01	MDACD5	692	0.1355 -		ISM02.4	PPB	3/28/2017 16:45	SMPL
I2407-02	MDACD6	294	0.0521 -		ISM02.4	PPB	3/28/2017 16:47	SMPL
I2407-03	MDACD8	333	0.0603 -		ISM02.4	PPB	3/28/2017 16:50	SMPL
I2407-04	MDACD8D	346	0.063 -		ISM02.4	PPB	3/28/2017 16:52	SMPL
I2407-05	MDACD8S	12046	2.513 -		ISM02.4	PPB	3/28/2017 16:54	SMPL
I2407-06	MDACD9	164	0.0249 -		ISM02.4	PPB	3/28/2017 16:56	SMPL
I2407-07	MDACE0	495	0.0942 -		ISM02.4	PPB	3/28/2017 16:59	SMPL
I2407-08	MDACE1	396	0.0735 -		ISM02.4	PPB	3/28/2017 17:01	SMPL
I2407-09	MDACE2	641	0.1248 -		ISM02.4	PPB	3/28/2017 17:03	SMPL
I2407-10	MDACE3	283	0.0498 -		ISM02.4	PPB	3/28/2017 17:06	SMPL
I2407-11	MDACE4	456	0.0861 -		ISM02.4	PPB	3/28/2017 17:08	SMPL
I2407-12	MDACE5	326	0.0588 -		ISM02.4	PPB	3/28/2017 17:10	SMPL
I2407-13	MDACE6	498	0.0949 -		ISM02.4	PPB	3/28/2017 17:12	SMPL
I2407-14	MDACC1	438036	HIGH -		ISM02.4	PPB	3/28/2017 17:15	

LB86587

INSTRUMENT ID : CV1

CCV096	CCV096	23520	4.9156 -	ISM02.4	PPB	3/28/2017 17:25	SMPL
CCB096	CCB096	-90	-0.0283 -	ISM02.4	PPB	3/28/2017 17:27	SMPL
I2407-14DLX10	MDACC1	68817	14.4007 -	ISM02.4	PPB	3/28/2017 17:30	SMPL
I2407-14DL2X50	MDACC1	13699	2.8591 -	ISM02.4	PPB	3/28/2017 17:46	SMPL
CCV097	CCV097	23663	4.9456 -	ISM02.4	PPB	3/28/2017 17:55	SMPL
CCB097	CCB097	-60	-0.022 -	ISM02.4	PPB	3/28/2017 17:57	SMPL