



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86588

Temp.: 22.5 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/28/2017	16:10
2	CAL2	1	Water			7.00	03/28/2017	16:11
3	CAL3	1	Water			10.03	03/28/2017	16:12
4	ICV	1	Water			7.01	03/28/2017	16:13
5	CCV1	1	Water			2.02	03/28/2017	16:15
6	I2435-02	1	Solid	20.03	20	8.84	03/28/2017	16:16
7	I2435-02DUP	1	Solid	20.05	20	8.85	03/28/2017	16:17
8	I2440-03	1	Solid	20.02	20	6.46	03/28/2017	16:18
9	I2440-06	1	Solid	20.04	20	7.70	03/28/2017	16:19
10	CCV2	1	Water			12.02	03/28/2017	16:20

1B86588

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2435 WorkList ID : 96919 Date : 3/28/2017 3:47:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2435-02	Corrosivity	Cool 4 deg C	URSC05	M53	HSP-01	03/27/2017	9045C
03/31/2017	Solid	I2440-03	Corrosivity	Cool 4 deg C	PSEG01	M51	ROLL-OFF-15-50	03/28/2017	9045C
03/31/2017	Solid	I2440-06	Corrosivity	Cool 4 deg C	PSEG01	M51	ROLL-OFF-20-41	03/28/2017	9045C

Date/Time 03/28/17 3:47PM
Received by: JR
Relinquished by: JR

Date/Time 03/28/17 6:30PM
Received by: JR
Relinquished by: JR