


7470A

LB86604
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34643



A= 0.0000e+000

B= 2.8850e-004

C= 1.3638e-002

Rho= 0.9999936

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	-39	0.000	-39				
0.2	0.200	0.193	-0.007	621	0.0 %	621				
2.5	2.500	2.502	0.002	8625	0.0 %	8625				
5.0	5.000	5.019	0.019	17348	0.0 %	17348				
7.5	7.500	7.476	-0.024	25866	0.0 %	25866				
10.0	10.000	10.008	0.008	34643	0.0 %	34643				

LB86604 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Mar 2017 11:48:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-39	-	0.0000	7470A	PPB	29 Mar 2017 12:15:22	S	Std
0.2 - 1	50.2	621	-	0.2000	7470A	PPB	29 Mar 2017 12:17:30	S	Std
2.5 - 1	52.5	8625	-	2.5000	7470A	PPB	29 Mar 2017 12:19:39	S	Std
5.0 - 1	55	17348	-	5.0000	7470A	PPB	29 Mar 2017 12:21:47	S	Std
7.5 - 1	57.5	25866	-	7.5000	7470A	PPB	29 Mar 2017 12:23:55	S	Std
10.0 - 1	511	34643	-	10.0000	7470A	PPB	29 Mar 2017 12:28:29	S	Std
ICV94 - 1	ICV94	13444	3.8923	-7470A	PPB	29 Mar 2017 12:30:59	U	SMPL	
ICB94 - 1	ICB94	-141	-0.0270	-7470A	PPB	29 Mar 2017 12:33:05	U	SMPL	
CCV01 - 1	CCV01	17517	5.0673	-7470A	PPB	29 Mar 2017 12:35:11	U	SMPL	
CCB01 - 1	CCB01	-130	-0.0239	-7470A	PPB	29 Mar 2017 12:37:18	U	SMPL	
CRA - 1	CRA	625	0.1940	-7470A	PPB	29 Mar 2017 12:39:24	U	SMPL	
HighStd - 1	HighStd	34060	9.8400	-7470A	PPB	29 Mar 2017 12:41:31	U	SMPL	
ChkStd - 1	ChkStd	24307	7.0263	-7470A	PPB	29 Mar 2017 12:43:38	U	SMPL	
PB97917BL - 1	PBW	-187	-0.0403	-7470A	PPB	29 Mar 2017 12:45:46	U	SMPL	
PB97917BS - 1	LCSW	14042	4.0648	-7470A	PPB	29 Mar 2017 12:47:54	U	SMPL	
PB97917TB - 1	PB97842TB	-118	-0.0204	-7470A	PPB	29 Mar 2017 12:50:02	U	SMPL	
I2421-02 - 1	HHD-GW1	-8	0.0113	-7470A	PPB	29 Mar 2017 12:52:10	U	SMPL	
I2421-04 - 1	HHD-GW2	-3	0.0128	-7470A	PPB	29 Mar 2017 12:54:18	U	SMPL	
I2422-02 - 1	HHD-S1	12	0.0171	-7470A	PPB	29 Mar 2017 12:56:27	U	SMPL	
I2422-02DUP - 1	HHD-S1DUP	10	0.0165	-7470A	PPB	29 Mar 2017 12:58:33	U	SMPL	
CCV02 - 1	CCV02	17271	4.9964	-7470A	PPB	29 Mar 2017 13:00:40	U	SMPL	
CCB02 - 1	CCB02	-141	-0.0270	-7470A	PPB	29 Mar 2017 13:02:46	U	SMPL	
I2422-02MS - 1	HHD-S1MS	14025	4.0599	-7470A	PPB	29 Mar 2017 13:04:53	U	SMPL	
I2422-02MSD - 1	HHD-S1MSD	13261	3.8395	-7470A	PPB	29 Mar 2017 13:07:00	U	SMPL	
PB97918BL - 1	PBW	-63	-0.0045	-7470A	PPB	29 Mar 2017 13:09:07	U	SMPL	
PB97918BS - 1	LCSW	14181	4.1049	-7470A	PPB	29 Mar 2017 13:11:14	U	SMPL	
I2296-05 - 1	LOD-WATER-05-QT2-2017	114	0.0465	-7470A	PPB	29 Mar 2017 13:13:22	U	SMPL	
I2296-06 - 1	LOQ-WATER-06-QT2-2017	1457	0.4340	-7470A	PPB	29 Mar 2017 13:15:29	U	SMPL	
I2422-02LX5 - 1		-26	0.0061	-7470A	PPB	29 Mar 2017 13:17:37	U	SMPL	
I2422-02A - 1		13919	4.0293	-7470A	PPB	29 Mar 2017 13:19:45	U	SMPL	
CCV03 - 1	CCV03	17543	5.0748	-7470A	PPB	29 Mar 2017 13:21:54	U	SMPL	
CCB03 - 1	CCB03	-144	-0.0279	-7470A	PPB	29 Mar 2017 13:24:01	U	SMPL	