



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86608

Temp.: 15.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/29/2017	13:10
2	CAL2	1	Water			7.01	03/29/2017	13:11
3	CAL3	1	Water			10.02	03/29/2017	13:12
4	ICV	1	Water			6.99	03/29/2017	13:13
5	CCV1	1	Water			2.01	03/29/2017	13:15
6	I2449-01	1	Water			5.46	03/29/2017	13:16
7	I2449-01DUP	1	Water			5.47	03/29/2017	13:17
8	CCV2	1	Water			12.02	03/29/2017	13:18

1786008

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2449 WorkList ID : 96953 Date : 3/29/2017 12:58:00 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2449-01-D	pH	Cool 4 deg C	STER01	M63	POOLWATER	03/28/2017	SM 4500-PH B

Date/Time 03/29/17 12:58 PM
Received by: JV
Relinquished by: [Signature]

Date/Time 03/29/17 21:00 PM
Received by: [Signature]
Relinquished by: JO