

245.1


L1386613

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33780



A= 0.0000e+000

B= 2.9736e-004

C= -8.2972e-003

Rho= 0.9999703

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	23	0.000	23				
0.2	0.200	0.202	0.002	707	0.0 %	707				
2.5	2.500	2.518	0.018	8496	0.0 %	8496				
5.0	5.000	5.001	0.001	16846	0.0 %	16846				
7.5	7.500	7.444	-0.056	25060	0.0 %	25060				
10.0	10.000	10.037	0.037	33780	0.0 %	33780				

LB86613 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Mar 2017 14:50:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	245.1	PPB	29 Mar 2017 14:52:09	S	Std
0.2 - 1	50.2	707	-	0.2000	245.1	PPB	29 Mar 2017 14:54:17	S	Std
2.5 - 1	52.5	8496	-	2.5000	245.1	PPB	29 Mar 2017 14:56:25	S	Std
5.0 - 1	55	16846	-	5.0000	245.1	PPB	29 Mar 2017 14:58:34	S	Std
7.5 - 1	57.5	25060	-	7.5000	245.1	PPB	29 Mar 2017 15:00:41	S	Std
10.0 - 1	510	33780	-	10.0000	245.1	PPB	29 Mar 2017 15:02:49	S	Std
ICV95 - 1	ICV95	13119	3.8928	-245.1		PPB	29 Mar 2017 15:05:50	U	SMPL
ICB95 - 1	ICB95	-38	-0.0196	-245.1		PPB	29 Mar 2017 15:07:56	U	SMPL
CCV04 - 1	CCV04	16707	4.9598	-245.1		PPB	29 Mar 2017 15:10:02	U	SMPL
CCB04 - 1	CCB04	-39	-0.0199	-245.1		PPB	29 Mar 2017 15:12:09	U	SMPL
CRA - 1	CRA	680	0.1939	-245.1		PPB	29 Mar 2017 15:14:16	U	SMPL
HighStd - 1	HighStd	33265	9.8835	-245.1		PPB	29 Mar 2017 15:16:22	U	SMPL
ChkStd - 1	ChkStd	23399	6.9497	-245.1		PPB	29 Mar 2017 15:18:29	U	SMPL
PB97919BL - 1	PBW	-69	-0.0288	-245.1		PPB	29 Mar 2017 15:20:37	U	SMPL
PB97919BS - 1	LCSW	14267	4.2342	-245.1		PPB	29 Mar 2017 15:22:46	U	SMPL
I2296-07 - 1	LOD-WATER-07-QT2-2017	193	0.0491	-245.1		PPB	29 Mar 2017 15:24:53	U	SMPL
I2296-08 - 1	LOQ-WATER-08-QT2-2017	1499	0.4375	-245.1		PPB	29 Mar 2017 15:27:02	U	SMPL
CCV05 - 1	CCV05	16414	4.8726	-245.1		PPB	29 Mar 2017 15:29:10	U	SMPL
CCB05 - 1	CCB05	-47	-0.0223	-245.1		PPB	29 Mar 2017 15:31:19	U	SMPL