


7471A

LB86614
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45319



A= 0.0000e+000

B= 2.2013e-004

C= -1.9361e-003

Rho= 0.9999538

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	15	0.000	15				
0.2	0.200	0.195	-0.005	895	0.0 %	895				
2.5	2.500	2.458	-0.042	11175	0.0 %	11175				
5.0	5.000	5.071	0.071	23047	0.0 %	23047				
7.5	7.500	7.500	0.000	34081	0.0 %	34081				
10.0	10.000	9.974	-0.026	45319	0.0 %	45319				

LB86614 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 29 Mar 2017 15:40:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.00007471A	PPB		29 Mar 2017 15:41:54	S	Std
0.2 - 1	50.2	895	-	0.20007471A	PPB		29 Mar 2017 15:44:11	S	Std
2.5 - 1	52.5	11175	-	2.50007471A	PPB		29 Mar 2017 15:46:29	S	Std
5.0 - 1	55	23047	-	5.00007471A	PPB		29 Mar 2017 15:48:47	S	Std
7.5 - 1	57.5	34081	-	7.50007471A	PPB		29 Mar 2017 15:51:04	S	Std
10.0 - 1	510	45319	-	10.00007471A	PPB		29 Mar 2017 15:56:35	S	Std
ICV96 - 1	ICV96	17941	3.9474	-7471A	PPB		29 Mar 2017 15:59:14	U	SMPL
ICB96 - 1	ICB96	-104	-0.0248	-7471A	PPB		29 Mar 2017 16:01:29	U	SMPL
CCV06 - 1	CCV06	21890	4.8166	-7471A	PPB		29 Mar 2017 16:03:45	U	SMPL
CCB06 - 1	CCB06	-104	-0.0248	-7471A	PPB		29 Mar 2017 16:06:01	U	SMPL
CRA - 1	CRA	902	0.1966	-7471A	PPB		29 Mar 2017 16:08:17	U	SMPL
HighStd - 1	HighStd	45856	10.0922	-7471A	PPB		29 Mar 2017 16:10:33	U	SMPL
ChkStd - 1	ChkStd	32185	7.0829	-7471A	PPB		29 Mar 2017 16:12:49	U	SMPL
PB97913BL - 1	PBS	-129	-0.0303	-7471A	PPB		29 Mar 2017 16:15:07	U	SMPL
PB97913BS - 1	LCSS	18371	4.0420	-7471A	PPB		29 Mar 2017 16:17:24	U	SMPL
I2296-01 - 1	LOD-SOIL-01-QT2-2017	262	0.0557	-7471A	PPB		29 Mar 2017 16:22:14	U	SMPL
I2296-02 - 1	LOQ-SOIL-02-QT2-2017	1845	0.4042	-7471A	PPB		29 Mar 2017 16:24:31	U	SMPL
I2436-02 - 1	1A-9-11	1498	0.3278	-7471A	PPB		29 Mar 2017 16:26:48	U	SMPL
I2436-06 - 1	54C-9-11	623	0.1352	-7471A	PPB		29 Mar 2017 16:29:06	U	SMPL
I2438-01 - 1	Q811-01	2876	0.6311	-7471A	PPB		29 Mar 2017 16:31:22	U	SMPL
CCV07 - 1	CCV07	21871	4.8125	-7471A	PPB		29 Mar 2017 16:33:38	U	SMPL
CCB07 - 1	CCB07	-39	-0.0105	-7471A	PPB		29 Mar 2017 16:35:54	U	SMPL
I2438-02 - 1	Q811-02	3227	0.7084	-7471A	PPB		29 Mar 2017 16:38:10	U	SMPL
I2438-03 - 1	Q811-03	5943	1.3063	-7471A	PPB		29 Mar 2017 16:40:26	U	SMPL
I2438-04 - 1	Q811-04	3136	0.6884	-7471A	PPB		29 Mar 2017 16:42:43	U	SMPL
I2438-05 - 1	Q811-05	3259	0.7155	-7471A	PPB		29 Mar 2017 16:44:59	U	SMPL
I2438-06 - 1	Q811-06	4462	0.9803	-7471A	PPB		29 Mar 2017 16:47:16	U	SMPL
I2438-07 - 1	Q811-07	3340	0.7333	-7471A	PPB		29 Mar 2017 16:49:33	U	SMPL
I2438-08 - 1	Q811-08	2772	0.6083	-7471A	PPB		29 Mar 2017 16:51:51	U	SMPL
I2438-09 - 1	Q811-09	5084	1.1172	-7471A	PPB		29 Mar 2017 16:54:08	U	SMPL
I2438-10 - 1	Q811-10	3325	0.7300	-7471A	PPB		29 Mar 2017 16:56:26	U	SMPL
I2438-11 - 1	Q811-11	3869	0.8497	-7471A	PPB		29 Mar 2017 16:58:42	U	SMPL
CCV08 - 1	CCV08	22224	4.8902	-7471A	PPB		29 Mar 2017 17:00:59	U	SMPL
CCB08 - 1	CCB08	-35	-0.0096	-7471A	PPB		29 Mar 2017 17:03:15	U	SMPL
I2440-01 - 1	ROLL-OFF-15-50	1731	0.3791	-7471A	PPB		29 Mar 2017 17:05:31	U	SMPL
I2440-01DUP - 1	ROLL-OFF-15-50DUP	1566	0.3428	-7471A	PPB		29 Mar 2017 17:07:48	U	SMPL
I2440-01MS - 1	ROLL-OFF-15-50MS	18111	3.9848	-7471A	PPB		29 Mar 2017 17:10:04	U	SMPL
I2440-01MSD - 1	ROLL-OFF-15-50MSD	18111	3.9848	-7471A	PPB		29 Mar 2017 17:12:21	U	SMPL
I2440-04 - 1	ROLL-OFF-20-41	493	0.1066	-7471A	PPB		29 Mar 2017 17:14:37	U	SMPL
PB97914BL - 1	PBS	12	0.0007	-7471A	PPB		29 Mar 2017 17:16:54	U	SMPL
PB97914BS - 1	LCSS	15763	3.4679	-7471A	PPB		29 Mar 2017 17:19:11	U	SMPL
I2443-01 - 1	E2-S11-BASE	67761	14.9141	-7471A	PPB		29 Mar 2017 17:21:29	U	SMPL
I2443-02 - 1	E2-Q14-BASE	2593	0.5689	-7471A	PPB		29 Mar 2017 17:23:47	U	SMPL
I2443-03 - 1	E2-P12-BASE	84103	18.5114	-7471A	PPB		29 Mar 2017 17:26:04	U	SMPL
CCV09 - 1	CCV09	22072	4.8567	-7471A	PPB		29 Mar 2017 17:28:20	U	SMPL
CCB09 - 1	CCB09	-57	-0.0145	-7471A	PPB		29 Mar 2017 17:30:37	U	SMPL
I2443-04 - 1	E2-P12.5-ESW	630	0.1367	-7471A	PPB		29 Mar 2017 17:32:54	U	SMPL
I2443-05 - 1	E2-O12-ESW	720	0.1566	-7471A	PPB		29 Mar 2017 17:35:10	U	SMPL
I2443-06 - 1	E2-BB10-BASE	19324	4.2518	-7471A	PPB		29 Mar 2017 17:37:26	U	SMPL
I2443-07 - 1	E2-BB10-ESW	17384	3.8248	-7471A	PPB		29 Mar 2017 17:39:42	U	SMPL
I2443-08 - 1	E2-BB9-BASE	11001	2.4197	-7471A	PPB		29 Mar 2017 17:41:59	U	SMPL
I2443-09 - 1	E2-AA9-BASE	1062	0.2318	-7471A	PPB		29 Mar 2017 17:44:16	U	SMPL
I2443-09DUP - 1	E2-AA9-BASEDUP	975	0.2127	-7471A	PPB		29 Mar 2017 17:46:33	U	SMPL
I2443-09MS - 1	E2-AA9-BASEMS	15847	3.4864	-7471A	PPB		29 Mar 2017 17:48:51	U	SMPL
I2443-09MSD - 1	E2-AA9-BASEMSD	16128	3.5483	-7471A	PPB		29 Mar 2017 17:51:09	U	SMPL
I2443-10 - 1	E2-AA9-ESW	11296	2.4846	-7471A	PPB		29 Mar 2017 17:53:25	U	SMPL
CCV10 - 1	CCV10	23247	5.1154	-7471A	PPB		29 Mar 2017 17:55:42	U	SMPL
CCB10 - 1	CCB10	-26	-0.0077	-7471A	PPB		29 Mar 2017 17:57:58	U	SMPL
I2443-11 - 1	E2-AA9-SSW	20491	4.5087	-7471A	PPB		29 Mar 2017 18:00:14	U	SMPL
I2443-12 - 1	E2-AA9-WSW	54984	12.1015	-7471A	PPB		29 Mar 2017 18:02:31	U	SMPL
I2443-13 - 1	E2-AA9-NSW	106148	23.3641	-7471A	PPB		29 Mar 2017 18:04:47	U	SMPL
I2443-14 - 1	E2-BB-CC-8-9-BASE	58948	12.9741	-7471A	PPB		29 Mar 2017 18:07:04	U	SMPL
I2443-15 - 1	E2-BB-CC-7-8-BASE	9690	2.1311	-7471A	PPB		29 Mar 2017 18:09:22	U	SMPL
I2440-01LX5 - 1		272	0.0579	-7471A	PPB		29 Mar 2017 18:11:39	U	SMPL
I2440-01A - 1		19522	4.2954	-7471A	PPB		29 Mar 2017 18:13:56	U	SMPL
I2443-09LX5 - 1		206	0.0434	-7471A	PPB		29 Mar 2017 18:16:14	U	SMPL
I2443-09A - 1		17971	3.9540	-7471A	PPB		29 Mar 2017 18:18:32	U	SMPL
CCV11 - 1	CCV11	23395	5.1479	-7471A	PPB		29 Mar 2017 18:20:49	U	SMPL
CCB11 - 1	CCB11	-36	-0.0099	-7471A	PPB		29 Mar 2017 18:23:05	U	SMPL
I2443-01DLX5 - 1		12885	2.8344	-7471A	PPB		29 Mar 2017 18:25:22	U	SMPL
I2443-03DLX5 - 1		15497	3.4094	-7471A	PPB		29 Mar 2017 18:27:39	U	SMPL
I2443-12DLX5 - 1		10977	2.4144	-7471A	PPB		29 Mar 2017 18:29:56	U	SMPL
I2443-13DLX5 - 1		21717	4.7786	-7471A	PPB		29 Mar 2017 18:32:12	U	SMPL
I2443-14DLX5 - 1		11652	2.5630	-7471A	PPB		29 Mar 2017 18:34:29	U	SMPL
CCV12 - 1	CCV12	22982	5.0570	-7471A	PPB		29 Mar 2017 18:36:46	U	SMPL
CCB12 - 1	CCB12	-22	-0.0068	-7471A	PPB		29 Mar 2017 18:39:02	U	SMPL