


7470A

LB86631  
INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

33799



A= 0.0000e+000

B= 2.9663e-004

C= 1.5501e-002

Rho= 0.9999428

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.043  | 0.043  | 93    | 0.000      | 93    |       |       |       |       |
| 0.2    | 0.200  | 0.225  | 0.025  | 706   | 0.0 %      | 706   |       |       |       |       |
| 2.5    | 2.500  | 2.437  | -0.063 | 8163  | 0.0 %      | 8163  |       |       |       |       |
| 5.0    | 5.000  | 4.969  | -0.031 | 16700 | 0.0 %      | 16700 |       |       |       |       |
| 7.5    | 7.500  | 7.485  | -0.015 | 25181 | 0.0 %      | 25181 |       |       |       |       |
| 10.0   | 10.000 | 10.041 | 0.041  | 33799 | 0.0 %      | 33799 |       |       |       |       |

## LB86631 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Mar 2017 11:40:03

| Sample ID       | Extended ID    | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|----------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50             | 93     | -       | 0.0000   | 7470A  | PPB   | 30 Mar 2017 11:47:58 | S    | Std  |
| 0.2 - 1         | 50.2           | 706    | -       | 0.2000   | 7470A  | PPB   | 30 Mar 2017 11:50:07 | S    | Std  |
| 2.5 - 1         | 52.5           | 8163   | -       | 2.5000   | 7470A  | PPB   | 30 Mar 2017 11:52:15 | S    | Std  |
| 5.0 - 1         | 55             | 16700  | -       | 5.0000   | 7470A  | PPB   | 30 Mar 2017 11:54:23 | S    | Std  |
| 7.5 - 1         | 57.5           | 25181  | -       | 7.5000   | 7470A  | PPB   | 30 Mar 2017 11:56:31 | S    | Std  |
| 10.0 - 1        | 511            | 33799  | -       | 10.0000  | 7470A  | PPB   | 30 Mar 2017 12:01:09 | S    | Std  |
| ICV97 - 1       | ICV97          | 13804  | 4.1101  | -        | 7470A  | PPB   | 30 Mar 2017 12:03:54 | U    | SMPL |
| ICB97 - 1       | ICB97          | -54    | -0.0005 | -        | 7470A  | PPB   | 30 Mar 2017 12:06:00 | U    | SMPL |
| CCV13 - 1       | CCV13          | 16999  | 5.0578  | -        | 7470A  | PPB   | 30 Mar 2017 12:08:06 | U    | SMPL |
| CCB13 - 1       | CCB13          | -47    | 0.0016  | -        | 7470A  | PPB   | 30 Mar 2017 12:10:13 | U    | SMPL |
| CRA - 1         | CRA            | 722    | 0.2297  | -        | 7470A  | PPB   | 30 Mar 2017 12:12:20 | U    | SMPL |
| HighStd - 1     | HighStd        | 33631  | 9.9913  | -        | 7470A  | PPB   | 30 Mar 2017 12:14:26 | U    | SMPL |
| ChkStd - 1      | ChkStd         | 24716  | 7.3469  | -        | 7470A  | PPB   | 30 Mar 2017 12:16:33 | U    | SMPL |
| PB97924BL - 1   | PBW            | -50    | 0.0007  | -        | 7470A  | PPB   | 30 Mar 2017 12:18:41 | U    | SMPL |
| PB97924BS - 1   | LCSW           | 13572  | 4.0413  | -        | 7470A  | PPB   | 30 Mar 2017 12:20:50 | U    | SMPL |
| PB97924TB - 1   | PB97891TB      | -8     | 0.0131  | -        | 7470A  | PPB   | 30 Mar 2017 12:22:57 | U    | SMPL |
| I2435-02 - 1    | HSP-01         | 63     | 0.0342  | -        | 7470A  | PPB   | 30 Mar 2017 12:25:06 | U    | SMPL |
| I2435-02DUP - 1 | HSP-01DUP      | 76     | 0.0380  | -        | 7470A  | PPB   | 30 Mar 2017 12:27:14 | U    | SMPL |
| I2435-02MS - 1  | HSP-01MS       | 12878  | 3.8354  | -        | 7470A  | PPB   | 30 Mar 2017 12:29:22 | U    | SMPL |
| I2435-02MSD - 1 | HSP-01MSD      | 12396  | 3.6925  | -        | 7470A  | PPB   | 30 Mar 2017 12:31:29 | U    | SMPL |
| CCV14 - 1       | CCV14          | 17388  | 5.1732  | -        | 7470A  | PPB   | 30 Mar 2017 12:33:35 | U    | SMPL |
| CCB14 - 1       | CCB14          | -42    | 0.0030  | -        | 7470A  | PPB   | 30 Mar 2017 12:35:41 | U    | SMPL |
| I2440-03 - 1    | ROLL-OFF-15-50 | 156    | 0.0618  | -        | 7470A  | PPB   | 30 Mar 2017 12:37:48 | U    | SMPL |
| I2440-06 - 1    | ROLL-OFF-20-41 | 191    | 0.0722  | -        | 7470A  | PPB   | 30 Mar 2017 12:39:54 | U    | SMPL |
| PB97925BL - 1   | PBW            | 179    | 0.0686  | -        | 7470A  | PPB   | 30 Mar 2017 12:42:01 | U    | SMPL |
| PB97925BS - 1   | LCSW           | 12818  | 3.8176  | -        | 7470A  | PPB   | 30 Mar 2017 12:44:08 | U    | SMPL |
| PB97925TB - 1   | PB97893TB      | -78    | -0.0076 | -        | 7470A  | PPB   | 30 Mar 2017 12:46:16 | U    | SMPL |
| I2436-04 - 1    | 1A-9-11        | 155    | 0.0615  | -        | 7470A  | PPB   | 30 Mar 2017 12:48:23 | U    | SMPL |
| I2436-04DUP - 1 | 1A-9-11DUP     | 156    | 0.0618  | -        | 7470A  | PPB   | 30 Mar 2017 12:50:31 | U    | SMPL |
| I2436-04MS - 1  | 1A-9-11MS      | 14684  | 4.3712  | -        | 7470A  | PPB   | 30 Mar 2017 12:52:39 | U    | SMPL |
| I2436-04MSD - 1 | 1A-9-11MSD     | 15000  | 4.4649  | -        | 7470A  | PPB   | 30 Mar 2017 12:54:48 | U    | SMPL |
| I2436-08 - 1    | 54C-9-11       | 1270   | 0.3922  | -        | 7470A  | PPB   | 30 Mar 2017 12:56:54 | U    | SMPL |
| CCV15 - 1       | CCV15          | 16619  | 4.9451  | -        | 7470A  | PPB   | 30 Mar 2017 12:59:01 | U    | SMPL |
| CCB15 - 1       | CCB15          | -24    | 0.0084  | -        | 7470A  | PPB   | 30 Mar 2017 13:01:07 | U    | SMPL |
| I2435-02LX5 - 1 |                | 83     | 0.0401  | -        | 7470A  | PPB   | 30 Mar 2017 13:03:14 | U    | SMPL |
| I2435-02A - 1   |                | 13664  | 4.0686  | -        | 7470A  | PPB   | 30 Mar 2017 13:05:21 | U    | SMPL |
| I2436-04LX5 - 1 |                | -2     | 0.0149  | -        | 7470A  | PPB   | 30 Mar 2017 13:07:27 | U    | SMPL |
| I2436-04A - 1   |                | 13844  | 4.1220  | -        | 7470A  | PPB   | 30 Mar 2017 13:09:34 | U    | SMPL |
| CCV16 - 1       | CCV16          | 17050  | 5.0730  | -        | 7470A  | PPB   | 30 Mar 2017 13:11:41 | U    | SMPL |
| CCB16 - 1       | CCB16          | -38    | 0.0042  | -        | 7470A  | PPB   | 30 Mar 2017 13:13:48 | U    | SMPL |