



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: JIGNESHDate: 03/30/2017Run Number: LB86641BalanceID: WC SC-3DATE & TIME IN: 03/30/2017 16:30DATE & TIME OUT: 03/30/2017 20:35OVEN TEMP: 104 °COVEN TEMP: 105 °C

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @105°C (g)	Weight (g)	Result mg/L
1	LB86641BL	LB86641BL	1.3025	1.3025	100	1.3025	1.3025	1.3025	0	0
2	LB86641BS	LB86641BS	1.4681	1.4681	100	1.523	1.523	1.523	0.0549	549
3	I2459-01	STORMDRAIN-1	1.4171	1.4171	800	1.4173	1.4173	1.4173	0.0002	0.3
4	I2459-02	STORMDRAIN-2	1.4289	1.4289	400	1.429	1.429	1.429	0.0001	0.3
5	I2459-02DUP	STORMDRAIN-2DUP	1.3961	1.3961	400	1.3963	1.3963	1.3963	0.0002	0.5
6	I2465-01	EFF-170329-04	1.4122	1.4122	800	1.4144	1.4144	1.4144	0.0022	2.8
7	I2471-01	SPDES-1	1.4183	1.4183	800	1.4185	1.4185	1.4185	0.0002	0.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS 12471

WorkList ID : 96984

Date : 3/30/2017 3:37:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2459-01-D	TSS	Cool 4 deg C	KWAR01	N13	STORMDRAIN-1	03/29/2017	SM2540 D
	Water	I2459-02-D	TSS	Cool 4 deg C	KWAR01	N13	STORMDRAIN-2	03/29/2017	SM2540 D
	Water	I2465-01-D	TSS	Cool 4 deg C	VIAS01	N21	EFF-170329-04	03/29/2017	SM2540 D
	Water	I2471-01-C	TSS	Cool 4 deg C	SKAN06	N21	SPDES-1	03/30/2017	SM2540 D

Date/Time 03/30/17 3:37 PM
 Received by:
 Relinquished by:

Date/Time 03/30/17 5:25 PM
 Received by:
 Relinquished by: