


245.1

LB 86650
INSTRUMENT ID: CVL

Linear

 μ Abs.:

31977



A= 0.0000e+000

B= 3.1243e-004

C= -3.1863e-003

Rho= 0.9999017

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	41	0.000	41				
0.2	0.200	0.202	0.002	656	0.0 %	656				
2.5	2.500	2.530	0.030	8107	0.0 %	8107				
5.0	5.000	4.900	-0.100	15695	0.0 %	15695				
7.5	7.500	7.571	0.071	24242	0.0 %	24242				
10.0	10.000	9.988	-0.012	31977	0.0 %	31977				

LB86650 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 31 Mar 2017 10:57:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	41	-	0.0000	245.1	PPB	31 Mar 2017 11:03:46	S	Std
0.2 - 1	50.2	656	-	0.2000	245.1	PPB	31 Mar 2017 11:05:54	S	Std
2.5 - 1	52.5	8107	-	2.5000	245.1	PPB	31 Mar 2017 11:08:02	S	Std
5.0 - 1	55	15695	-	5.0000	245.1	PPB	31 Mar 2017 11:10:10	S	Std
7.5 - 1	57.5	24242	-	7.5000	245.1	PPB	31 Mar 2017 11:12:18	S	Std
10.0 - 1	90	31977	-	10.0000	245.1	PPB	31 Mar 2017 11:14:26	S	Std
ICV98 - 1	ICV98	12306	3.8416	-245.1		PPB	31 Mar 2017 11:17:04	U	SMPL
ICB98 - 1	ICB98	-25	-0.0110	-245.1		PPB	31 Mar 2017 11:19:10	U	SMPL
CCV17 - 1	CCV17	15982	4.9901	-245.1		PPB	31 Mar 2017 11:21:16	U	SMPL
CCB17 - 1	CCB17	-50	-0.0188	-245.1		PPB	31 Mar 2017 11:23:23	U	SMPL
CRA - 1	CRA	671	0.2065	-245.1		PPB	31 Mar 2017 11:25:30	U	SMPL
HighStd - 1	HighStd	31875	9.9557	-245.1		PPB	31 Mar 2017 11:27:37	U	SMPL
ChkStd - 1	ChkStd	22247	6.9476	-245.1		PPB	31 Mar 2017 11:29:44	U	SMPL
PB97941BL - 1	PBW	126	0.0362	-245.1		PPB	31 Mar 2017 11:31:53	U	SMPL
PB97941BS - 1	LCSW	12741	3.9775	-245.1		PPB	31 Mar 2017 11:34:02	U	SMPL
I2445-01 - 1	032717-PRRT-F-D-M	142	0.0412	-245.1		PPB	31 Mar 2017 11:38:36	U	SMPL
I2445-02 - 1	032717-PRRT-F-U-M	108	0.0306	-245.1		PPB	31 Mar 2017 11:43:07	U	SMPL
I2446-01 - 1	032817-PRRT-F-D-M	117	0.0334	-245.1		PPB	31 Mar 2017 11:47:46	U	SMPL
I2446-02 - 1	032817-PRRT-F-U-M	161	0.0471	-245.1		PPB	31 Mar 2017 11:49:54	U	SMPL
I2446-02DUP - 1	032817-PRRT-F-U-MDUP	79	0.0215	-245.1		PPB	31 Mar 2017 11:52:01	U	SMPL
CCV18 - 1	CCV18	16011	4.9992	-245.1		PPB	31 Mar 2017 11:56:56	U	SMPL
CCB18 - 1	CCB18	-37	-0.0147	-245.1		PPB	31 Mar 2017 11:59:03	U	SMPL
I2446-02MS - 1	032817-PRRT-F-U-MMS	15644	4.8845	-245.1		PPB	31 Mar 2017 12:01:10	U	SMPL
I2446-02MSD - 1	032817-PRRT-F-U-MMSD	16044	5.0095	-245.1		PPB	31 Mar 2017 12:03:17	U	SMPL
I2446-02LX5 - 1		-2	-0.0038	-245.1		PPB	31 Mar 2017 12:05:24	U	SMPL
I2446-02A - 1		12651	3.9494	-245.1		PPB	31 Mar 2017 12:07:31	U	SMPL
CCV19 - 1	CCV19	16058	5.0139	-245.1		PPB	31 Mar 2017 12:09:38	U	SMPL
CCB19 - 1	CCB19	-56	-0.0207	-245.1		PPB	31 Mar 2017 12:11:46	U	SMPL