



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86654

Temp.: 24.4 °C

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/31/2017	13:30
2	CAL2	1	Water			7.00	03/31/2017	13:31
3	CAL3	1	Water			10.02	03/31/2017	13:32
4	ICV	1	Water			7.00	03/31/2017	13:33
5	CCV1	1	Water			2.02	03/31/2017	13:34
6	I2491-01	1	Water			5.95	03/31/2017	13:35
7	I2491-01DUP	1	Water			5.96	03/31/2017	13:36
8	I2491-03	1	Water			6.65	03/31/2017	13:39
9	I2491-05	1	Water			6.54	03/31/2017	13:40
10	I2491-07	1	Water			6.56	03/31/2017	13:41
11	CCV2	1	Water			12.01	03/31/2017	13:42

LB8654

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12491

WorkList ID : 97010

Date : 3/31/2017 1:11:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	12491-01-1	pH	Cool 4 deg C	LOCK01	N31	MW-1	03/30/2017	SM 4500-PH B
	Water	12491-03-1	pH	Cool 4 deg C	LOCK01	N31	MW-2	03/30/2017	SM 4500-PH B
	Water	12491-05-1	pH	Cool 4 deg C	LOCK01	N31	MW-3	03/30/2017	SM 4500-PH B
	Water	12491-07-1	pH	Cool 4 deg C	LOCK01	N31	MW-4	03/30/2017	SM 4500-PH B

Date/Time 03/31/17 1:11 PM
Received by: CP
Relinquished by: _____

Date/Time 03/31/17 3:00 PM
Received by: CP
Relinquished by: JP