

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB86656

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP53917
calibration std. phosphate 0.5 ppm	WP53916
calibration std. phosphate 0.3 ppm	WP53915
calibration std. 0 ppm	WP53912
phosphate CCV std.	WP53918
calibration std. phosphate 0.1 ppm	WP53914
calibration std. phosphate 0.05 ppm	WP53913
Combined reagent	WP53921
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate ICV std.	WP53919

Intercept: -0.0009Slope: 0.6468Regression: 0.999814

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	03/31/2017	11:55
2	CAL2	0.05	1	50	50	0.032	03/31/2017	11:55
3	CAL3	0.10	1	50	50	0.065	03/31/2017	11:56
4	CAL4	0.30	1	50	50	0.185	03/31/2017	11:56
5	CAL5	0.50	1	50	50	0.329	03/31/2017	11:57
6	CAL6	1.00	1	50	50	0.645	03/31/2017	11:57



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB86656

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.316	0.490	03/31/2017	11:58
2	ICB		1	50	50	0.000	0.001	03/31/2017	11:58
3	CCV1	0.50	1	50	50	0.326	0.505	03/31/2017	11:59
4	CCB1		1	50	50	0.000	0.001	03/31/2017	11:59
5	PB97963BL		1	50	50	0.000	0.001	03/31/2017	12:00
6	PB97963BS	0.50	1	50	50	0.315	0.488	03/31/2017	12:00
7	I2459-01		1	50	50	0.011	0.018	03/31/2017	12:01
8	I2459-01DUP		1	50	50	0.013	0.021	03/31/2017	12:01
9	I2459-01MS	0.50	1	50	50	0.310	0.481	03/31/2017	12:02
10	I2459-01MSD	0.50	1	50	50	0.311	0.482	03/31/2017	12:02
11	I2459-02		1	50	50	0.021	0.034	03/31/2017	12:03
12	CCV2	0.50	1	50	50	0.320	0.496	03/31/2017	12:03
13	CCB2		1	50	50	0.000	0.001	03/31/2017	12:04