



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86663

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	03/31/2017	16:30
2	CAL2	1	Water			7.02	03/31/2017	16:31
3	CAL3	1	Water			10.04	03/31/2017	16:32
4	ICV	1	Water			7.01	03/31/2017	16:33
5	CCV1	1	Water			2.02	03/31/2017	16:34
6	I2488-02	1	Solid	20.02	20	8.48	03/31/2017	16:35
7	I2488-02DUP	1	Solid	20.04	20	8.49	03/31/2017	16:36
8	I2493-17	1	Solid	20.02	20	7.62	03/31/2017	16:37
9	I2493-19	1	Solid	20.01	20	7.64	03/31/2017	16:40
10	I2493-25	1	Solid	20.03	20	7.98	03/31/2017	16:41
11	I2493-27	1	Solid	20.05	20	7.97	03/31/2017	16:42
12	I2495-11	1	Solid	20.04	20	7.84	03/31/2017	16:43
13	I2495-12	1	Solid	20.01	20	7.69	03/31/2017	16:44
14	I2495-13	1	Solid	20.02	20	7.94	03/31/2017	16:45
15	I2495-14	1	Solid	20.04	20	7.57	03/31/2017	16:46
16	CCV2	1	Water			12.01	03/31/2017	16:48
17	I2500-03	1	Solid	20.05	20	8.59	03/31/2017	16:49
18	I2508-02	1	Solid	20.04	20	8.33	03/31/2017	16:50
19	I2508-04	1	Solid	20.02	20	6.78	03/31/2017	16:51
20	CCV3	1	Water			12.02	03/31/2017	16:52

LB 86663

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2488 WorkList ID : 97009 Date : 3/31/2017 1:09:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2488-02	Corrosivity	Cool 4 deg C	VIAS01	N31	WC-COMP-13	03/30/2017	9045C
	Solid	I2493-17	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP05-COMP	03/30/2017	9045C
	Solid	I2493-19	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP06-COMP	03/30/2017	9045C
	Solid	I2493-25	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP08-COMP	03/30/2017	9045C
	Solid	I2493-27	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP10-COMP	03/30/2017	9045C
	Solid	I2495-11	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP01-COMP	03/30/2017	9045C
	Solid	I2495-12	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP02-COMP	03/30/2017	9045C
	Solid	I2495-13	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP03-COMP	03/30/2017	9045C
	Solid	I2495-14	Corrosivity	Cool 4 deg C	CHAZ02	N51	LSS-SP04-COMP	03/30/2017	9045C
	Solid	I2500-03	Corrosivity	Cool 4 deg C	URSG01	N41	SS-2	03/31/2017	9045C
	Solid	I2508-02	Corrosivity	Cool 4 deg C	ENTA04	N13	WC-B58-B59-001	03/30/2017	9045C
	Solid	I2508-04	Corrosivity	Cool 4 deg C	ENTA04	N13	WC-B53-B57-001	03/31/2017	9045C

Date/Time 03/31/17 1:09 PM
 Received by: JH
 Relinquished by: CB

Date/Time 03/31/17 6:30 PM
 Received by: CD
 Relinquished by: SO