


7470A

LB 86666
INSTRUMENT ID: CV2

Linear

 μ Abs.:

34027



A= 0.0000e+000

B= 2.9379e-004

C= 3.2117e-002

Rho= 0.9999077

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.032	0.032	0	0.000	0				
0.2	0.200	0.228	0.028	668	0.0 %	668				
2.5	2.500	2.406	-0.094	8079	0.0 %	8079				
5.0	5.000	5.044	0.044	17061	0.0 %	17061				
7.5	7.500	7.461	-0.039	25286	0.0 %	25286				
10.0	10.000	10.029	0.029	34027	0.0 %	34027				

LB86666 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Apr 2017 09:51:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	0	-	0.0000	7470A	PPB	03 Apr 2017 09:57:28	S	Std
0.2 - 1	S0.2	668	-	0.2000	7470A	PPB	03 Apr 2017 09:59:36	S	Std
2.5 - 1	S2.5	8079	-	2.5000	7470A	PPB	03 Apr 2017 10:01:45	S	Std
5.0 - 1	S5	17061	-	5.0000	7470A	PPB	03 Apr 2017 10:03:53	S	Std
7.5 - 1	S7.5	25286	-	7.5000	7470A	PPB	03 Apr 2017 10:06:01	S	Std
10.0 - 1	S10	34027	-	10.0000	7470A	PPB	03 Apr 2017 10:10:35	S	Std
ICV01 - 1	ICV01	13707	4.0590	-7470A	PPB	PPB	03 Apr 2017 10:13:25	U	SMPL
ICB01 - 1	ICB01	-110	-0.0002	-7470A	PPB	PPB	03 Apr 2017 10:15:31	U	SMPL
CCV27 - 1	CCV27	17925	5.2982	-7470A	PPB	PPB	03 Apr 2017 10:17:38	U	SMPL
CCB27 - 1	CCB27	-127	-0.0052	-7470A	PPB	PPB	03 Apr 2017 10:19:44	U	SMPL
CRA - 1	CRA	635	0.2187	-7470A	PPB	PPB	03 Apr 2017 10:21:51	U	SMPL
HighStd - 1	HighStd	34605	10.1986	-7470A	PPB	PPB	03 Apr 2017 10:23:58	U	SMPL
ChkStd - 1	ChkStd	24106	7.1141	-7470A	PPB	PPB	03 Apr 2017 10:26:05	U	SMPL
PB97971BL - 1	PBW	-174	-0.0190	-7470A	PPB	PPB	03 Apr 2017 10:28:13	U	SMPL
PB97971BS - 1	LCSW	13956	4.1322	-7470A	PPB	PPB	03 Apr 2017 10:30:21	U	SMPL
PB97971TB - 1	PB97945TB	-122	-0.0037	-7470A	PPB	PPB	03 Apr 2017 10:32:29	U	SMPL
I2474-02 - 1	49B-9-11	161	0.0794	-7470A	PPB	PPB	03 Apr 2017 10:44:24	U	SMPL
I2474-02DUP - 1	49B-9-11DUP	163	0.0800	-7470A	PPB	PPB	03 Apr 2017 10:49:02	U	SMPL
I2474-02MS - 1	49B-9-11MS	15685	4.6401	-7470A	PPB	PPB	03 Apr 2017 10:51:10	U	SMPL
I2474-02MSD - 1	49B-9-11MSD	15232	4.5071	-7470A	PPB	PPB	03 Apr 2017 10:53:17	U	SMPL
CCV28 - 1	CCV28	18170	5.3702	-7470A	PPB	PPB	03 Apr 2017 10:55:23	U	SMPL
CCB28 - 1	CCB28	-65	0.0130	-7470A	PPB	PPB	03 Apr 2017 10:57:30	U	SMPL
I2474-02LX5 - 1		28	0.0403	-7470A	PPB	PPB	03 Apr 2017 10:59:36	U	SMPL
I2474-02A - 1		14175	4.1965	-7470A	PPB	PPB	03 Apr 2017 11:01:43	U	SMPL
CCV29 - 1	CCV29	18046	5.3338	-7470A	PPB	PPB	03 Apr 2017 11:03:50	U	SMPL
CCB29 - 1	CCB29	-79	0.0089	-7470A	PPB	PPB	03 Apr 2017 11:05:57	U	SMPL