


245.1 LB86670
INSTRUMENT ID: CV1

Linear

μ Abs.:

33355



A= 0.0000e+000

B= 2.9697e-004

C= 8.7703e-003

Rho= 0.9996941

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	8	0.000	8				
0.2	0.200	0.202	0.002	651	0.0 %	651				
2.5	2.500	2.500	-0.000	8388	0.0 %	8388				
5.0	5.000	4.895	-0.105	16454	0.0 %	16454				
7.5	7.500	7.678	0.178	25823	0.0 %	25823				
10.0	10.000	9.914	-0.086	33355	0.0 %	33355				

LB86670 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 03 Apr 2017 11:17:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	245.1	PPB	03 Apr 2017 11:21:38	S	Std
0.2 - 1	50.2	651	-	0.2000	245.1	PPB	03 Apr 2017 11:23:46	S	Std
2.5 - 1	52.5	8388	-	2.5000	245.1	PPB	03 Apr 2017 11:25:54	S	Std
5.0 - 1	55	16454	-	5.0000	245.1	PPB	03 Apr 2017 11:28:03	S	Std
7.5 - 1	57.5	25823	-	7.5000	245.1	PPB	03 Apr 2017 11:30:10	S	Std
10.0 - 1	510	33355	-	10.0000	245.1	PPB	03 Apr 2017 11:32:18	S	Std
ICV02 - 1	ICV02	13258	3.9460	-245.1		PPB	03 Apr 2017 11:34:55	U	SMPL
ICB02 - 1	ICB02	-78	-0.0144	-245.1		PPB	03 Apr 2017 11:37:01	U	SMPL
CCV30 - 1	CCV30	16409	4.8818	-245.1		PPB	03 Apr 2017 11:39:08	U	SMPL
CCB30 - 1	CCB30	-68	-0.0114	-245.1		PPB	03 Apr 2017 11:41:14	U	SMPL
CRA - 1	CRA	627	0.1950	-245.1		PPB	03 Apr 2017 11:43:22	U	SMPL
HighStd - 1	HighStd	33034	9.8190	-245.1		PPB	03 Apr 2017 11:45:28	U	SMPL
ChkStd - 1	ChkStd	23480	6.9817	-245.1		PPB	03 Apr 2017 11:47:35	U	SMPL
PB97972BL - 1	PBW	-128	-0.0292	-245.1		PPB	03 Apr 2017 11:49:44	U	SMPL
PB97972BS - 1	LCSW	14507	4.3170	-245.1		PPB	03 Apr 2017 11:51:52	U	SMPL
I2484-01 - 1	033017-PRRT-F-U-M	-4	0.0076	-245.1		PPB	03 Apr 2017 11:54:00	U	SMPL
I2484-02 - 1	033017-PRRT-F-D-M	26	0.0165	-245.1		PPB	03 Apr 2017 11:56:08	U	SMPL
I2485-01 - 1	032917-PRRT-F-U-M	72	0.0302	-245.1		PPB	03 Apr 2017 11:58:16	U	SMPL
I2485-02 - 1	032917-PRRT-F-D-M	63	0.0275	-245.1		PPB	03 Apr 2017 12:00:25	U	SMPL
I2485-03 - 1	032917-DSDC-F-D-M	56	0.0254	-245.1		PPB	03 Apr 2017 12:02:31	U	SMPL
CCV31 - 1	CCV31	16525	4.9162	-245.1		PPB	03 Apr 2017 12:04:38	U	SMPL
CCB31 - 1	CCB31	-44	-0.0043	-245.1		PPB	03 Apr 2017 12:06:45	U	SMPL
I2485-04 - 1	032917-DSDC-F-U-M	49	0.0233	-245.1		PPB	03 Apr 2017 12:08:52	U	SMPL
I2485-04DUP - 1	032917-DSDC-F-U-MDUP	37	0.0198	-245.1		PPB	03 Apr 2017 12:10:59	U	SMPL
I2485-04MS - 1	032917-DSDC-F-U-MMS	15868	4.7211	-245.1		PPB	03 Apr 2017 12:13:06	U	SMPL
I2485-04MSD - 1	032917-DSDC-F-U-MMSD	16735	4.9786	-245.1		PPB	03 Apr 2017 12:15:13	U	SMPL
I2485-04LX5 - 1		-36	-0.0019	-245.1		PPB	03 Apr 2017 12:17:20	U	SMPL
I2485-04A - 1		16617	4.9436	-245.1		PPB	03 Apr 2017 12:19:28	U	SMPL
CCV32 - 1	CCV32	16686	4.9641	-245.1		PPB	03 Apr 2017 12:21:36	U	SMPL
CCB32 - 1	CCB32	-49	-0.0058	-245.1		PPB	03 Apr 2017 12:23:44	U	SMPL