



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86682

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP53943
COD CCV std, 50ppm	WP53944

Temp In (C): <u>150</u>	Date In: <u>04/03/2017</u>	Time In: <u>13:20</u>
Temp Out (C): <u>150</u>	Date Out: <u>04/03/2017</u>	Time Out: <u>15:20</u>

Intercept: -0.5466

Slope: 1.0088

Regression: 1

Seq	Lab ID	True Value	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	03/16/2017	13:50
2	CAL2	10	1	Water	9.000	03/16/2017	13:50
3	CAL3	50	1	Water	50.000	03/16/2017	13:51
4	CAL4	100	1	Water	100.000	03/16/2017	13:51
5	CAL5	150	1	Water	151.000	03/16/2017	13:52



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB86682

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	51.000	51.097	03/16/2017	13:52
2	ICB				1	Water	0.000	0.542	03/16/2017	13:53
3	CCV1	50			1	Water	53.000	53.080	04/03/2017	16:02
4	CCB1				1	Water	0.000	0.542	04/03/2017	16:02
5	LB86682BL				1	Water	0.000	0.542	04/03/2017	16:03
6	LB86682BS	50			1	Water	51.000	51.097	04/03/2017	16:03
7	I2502-01				1	Water	12.000	12.437	04/03/2017	16:04
8	I2502-01DUP				1	Water	10.000	10.455	04/03/2017	16:04
9	I2502-01MS	50			1	Water	63.000	62.992	04/03/2017	16:05
10	I2502-01MSD	50			1	Water	64.000	63.984	04/03/2017	16:05
11	CCV2	50			1	Water	54.000	54.071	04/03/2017	16:06
12	CCB2				1	Water	0.000	0.542	04/03/2017	16:06

WORKLIST(Hardcopy Internal Chain)

LB 86682

WorkList Name : COD040317

WorkList ID : 97059

Date : 4/3/2017 1:05:38 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2502-01	COD	Conc H2SO4 tk	ATCE02	N31	OF-1	03/31/2017	SM5220 D

Date/Time 04/03/17 13:08
 Received by: IB
 Relinquished by: CR

Date/Time 04/03/17
 Received by: CR
 Relinquished by: IB