



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86699

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/04/2017	13:45
2	CAL2	1	Water			7.00	04/04/2017	13:46
3	CAL3	1	Water			10.03	04/04/2017	13:47
4	ICV	1	Water			7.01	04/04/2017	13:49
5	CCV1	1	Water			2.02	04/04/2017	14:10
6	I2529-05	1	Solid	20.03	20	8.73	04/04/2017	14:11
7	I2529-05DUP	1	Solid	20.05	20	8.74	04/04/2017	14:12
8	I2529-06	1	Solid	20.01	20	8.77	04/04/2017	14:15
9	I2530-01	1	Solid	20.04	20	7.42	04/04/2017	14:16
10	I2530-02	1	Solid	20.00	20	7.48	04/04/2017	14:17
11	I2530-03	1	Solid	20.03	20	7.10	04/04/2017	14:20
12	I2530-04	1	Solid	20.05	20	6.63	04/04/2017	14:21
13	I2530-05	1	Solid	20.04	20	6.56	04/04/2017	14:22
14	I2530-06	1	Solid	20.01	20	6.58	04/04/2017	14:23
15	I2530-07	1	Solid	20.03	20	5.10	04/04/2017	14:24
16	CCV2	1	Water			12.02	04/04/2017	14:25
17	I2530-08	1	Solid	20.05	20	4.50	04/04/2017	14:27
18	I2530-09	1	Solid	20.02	20	4.38	04/04/2017	14:30
19	I2530-10	1	Solid	20.05	20	7.28	04/04/2017	14:32
20	I2530-11	1	Solid	20.03	20	7.32	04/04/2017	14:33

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I2530-12	1	Solid	20.02	20	7.17	04/04/2017	14:34
22	CCV3	1	Water			12.01	04/04/2017	14:35

2386699

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2530

WorkList ID : 97099

Date : 4/4/2017 1:36:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I2529-05 ✓	pH	Cool 4 deg C	ECHE01	N23	EASTEND	03/11/2017	9045C
	Solid	I2529-06 ✓	pH	Cool 4 deg C	ECHE01	N23	WESTEND	03/11/2017	9045C
	Solid	I2530-01 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-1-24-36	03/21/2017	9045C
	Solid	I2530-02 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-1-20-24	03/21/2017	9045C
	Solid	I2530-03 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-1-36-45	03/21/2017	9045C
	Solid	I2530-04 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-2-12-24	03/21/2017	9045C
	Solid	I2530-05 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-2-24-36	03/21/2017	9045C
	Solid	I2530-06 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-2-36-45	03/21/2017	9045C
	Solid	I2530-07 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-3-12-24	03/21/2017	9045C
	Solid	I2530-08 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-3-24-36	03/21/2017	9045C
	Solid	I2530-09 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-3-36-45	03/21/2017	9045C
	Solid	I2530-10 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-4-12-24	03/21/2017	9045C
	Solid	I2530-11 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-4-24-36	03/21/2017	9045C
	Solid	I2530-12 ✓	pH	Cool 4 deg C	ECHE01	N32	SB-4-36-45	03/21/2017	9045C

Date/Time 04/04/17 1:36 PM
 Received by: 31
 Relinquished by: J.C.

Date/Time 04/04/17 1:25 PM
 Received by: J.C.
 Relinquished by: JN