



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86700

Temp.: 22.1 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/04/2017	13:45
2	CAL2	1	Water			7.00	04/04/2017	13:46
3	CAL3	1	Water			10.03	04/04/2017	13:47
4	ICV	1	Water			7.01	04/04/2017	13:49
5	CCV1	1	Water			2.01	04/04/2017	13:50
6	I2529-01	1	Water			7.82	04/04/2017	13:51
7	I2529-02	1	Water			7.50	04/04/2017	13:52
8	I2529-03	1	Water			7.53	04/04/2017	13:53
9	I2529-04	1	Water			7.63	04/04/2017	13:54
10	I2531-01	1	Water			4.42	04/04/2017	13:55
11	I2531-01DUP	1	Water			4.43	04/04/2017	13:56
12	CCV2	1	Water			12.02	04/04/2017	14:00

2886700

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 12529

WorkList ID : 97098

Date : 4/4/2017 1:36:04 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I2529-01 -B	pH	Cool 4 deg C	ECHE01	N23	S1-JOINT3	03/11/2017	SM 4500-PH B
	Water	I2529-02 -B	pH	Cool 4 deg C	ECHE01	N23	S2-MIDSTREAM	03/11/2017	SM 4500-PH B
	Water	I2529-03 -B	pH	Cool 4 deg C	ECHE01	N23	S3-DOWNSTREAM	03/11/2017	SM 4500-PH B
	Water	I2529-04 -B	pH	Cool 4 deg C	ECHE01	N23	S4-UPSTREAM	03/11/2017	SM 4500-PH B
	Water	I2531-01 -A	pH	Cool 4 deg C	STER05	N32	POOLWATER	04/03/2017	SM 4500-PH B

Date/Time 04/04/17 11:38 AM
 Received by: 74
 Relinquished by: CD

Date/Time 04/04/17 4:00 PM
 Received by: CP
 Relinquished by: JP