



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86702

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/04/2017	18:00
2	CAL2	1	Water			7.02	04/04/2017	18:01
3	CAL3	1	Water			10.03	04/04/2017	18:02
4	ICV	1	Water			7.01	04/04/2017	18:03
5	CCV1	1	Water			2.02	04/04/2017	18:05
6	I2537-05	1	Solid	20.03	20	8.89	04/04/2017	18:06
7	I2537-05DUP	1	Solid	20.05	20	8.90	04/04/2017	18:07
8	I2537-10	1	Solid	20.02	20	7.28	04/04/2017	18:08
9	I2537-15	1	Solid	20.04	20	7.17	04/04/2017	18:10
10	I2537-20	1	Solid	20.01	20	7.93	04/04/2017	18:11
11	I2538-10	1	Solid	20.03	20	8.79	04/04/2017	18:13
12	CCV2	1	Water			12.02	04/04/2017	18:15

WORKLIST(Hardcopy Internal Chain)

1386702

WorkList Name : CORROSIONITY I2537

WorkList ID : 97118

Date : 4/4/2017 5:33:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/07/2017	Solid	I2537-05	Corrosivity	Cool 4 deg C	PSEG01	N43	B1-(0-5)	04/04/2017	9045C
04/07/2017	Solid	I2537-10	Corrosivity	Cool 4 deg C	PSEG01	N43	B4-(0-5)	04/04/2017	9045C
04/07/2017	Solid	I2537-15	Corrosivity	Cool 4 deg C	PSEG01	N43	B5-(0-5)	04/04/2017	9045C
04/07/2017	Solid	I2537-20	Corrosivity	Cool 4 deg C	PSEG01	N43	B8-(0-5)	04/04/2017	9045C
04/07/2017	Solid	I2538-10	Corrosivity	Cool 4 deg C	PSEG01	N43	LOCUST-ST-IP(3-5)	04/04/2017	9045C

Date/Time 04/04/17 5:33PM
 Received by: JH
 Relinquished by: JH

Date/Time 04/04/17 7:00 PM
 Received by: CH
 Relinquished by: JH