

LB 86706


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34219



A= 0.0000e+000

B= 2.9409e-004

C= 1.3080e-002

Rho= 0.9998855

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	74	0.000	74				
0.2	0.200	0.215	0.015	688	0.0 %	688				
2.5	2.500	2.462	-0.038	8328	0.0 %	8328				
5.0	5.000	5.010	0.010	16991	0.0 %	16991				
7.5	7.500	7.401	-0.099	25120	0.0 %	25120				
10.0	10.000	10.077	0.077	34219	0.0 %	34219				

LB86706 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Apr 2017 09:46:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.0000	245.1	PPB	05 Apr 2017 09:54:02	S	Std
0.2 - 1	50.2	688	-	0.2000	245.1	PPB	05 Apr 2017 09:56:10	S	Std
2.5 - 1	52.5	8328	-	2.5000	245.1	PPB	05 Apr 2017 09:58:18	S	Std
5.0 - 1	55	16991	-	5.0000	245.1	PPB	05 Apr 2017 10:00:27	S	Std
7.5 - 1	57.5	25120	-	7.5000	245.1	PPB	05 Apr 2017 10:05:00	S	Std
10.0 - 1	510	34219	-	10.0000	245.1	PPB	05 Apr 2017 10:07:07	S	Std
ICV05 - 1	ICV05	13333	3.9342	-245.1		PPB	05 Apr 2017 10:09:43	U	SMPL
ICB05 - 1	ICB05	-23	0.0063	-245.1		PPB	05 Apr 2017 10:11:49	U	SMPL
CCV49 - 1	CCV49	16989	5.0094	-245.1		PPB	05 Apr 2017 10:13:56	U	SMPL
CCB49 - 1	CCB49	-8	0.0107	-245.1		PPB	05 Apr 2017 10:16:02	U	SMPL
CRA - 1	CRA	674	0.2113	-245.1		PPB	05 Apr 2017 10:18:09	U	SMPL
HighStd - 1	HighStd	34318	10.1058	-245.1		PPB	05 Apr 2017 10:20:15	U	SMPL
ChkStd - 1	ChkStd	24306	7.1613	-245.1		PPB	05 Apr 2017 10:22:22	U	SMPL
PB98034BL - 1	PBW	-118	-0.0216	-245.1		PPB	05 Apr 2017 10:24:31	U	SMPL
PB98034BS - 1	LCSW	13288	3.9210	-245.1		PPB	05 Apr 2017 10:26:39	U	SMPL
I2518-01 - 1	2AVE-03	102	0.0431	-245.1		PPB	05 Apr 2017 10:33:29	U	SMPL
I2518-01DUP - 1	2AVE-03DUP	101	0.0428	-245.1		PPB	05 Apr 2017 10:38:10	U	SMPL
I2518-01MS - 1	2AVE-03MS	15054	4.4404	-245.1		PPB	05 Apr 2017 10:40:17	U	SMPL
I2518-01MSD - 1	2AVE-03MSD	15409	4.5448	-245.1		PPB	05 Apr 2017 10:42:25	U	SMPL
CCV50 - 1	CCV50	16910	4.9862	-245.1		PPB	05 Apr 2017 10:44:32	U	SMPL
CCB50 - 1	CCB50	-11	0.0098	-245.1		PPB	05 Apr 2017 10:46:39	U	SMPL
I2518-01LX5 - 1		54	0.0290	-245.1		PPB	05 Apr 2017 10:48:46	U	SMPL
I2518-01A - 1		13453	3.9695	-245.1		PPB	05 Apr 2017 10:50:52	U	SMPL
CCV51 - 1	CCV51	16905	4.9847	-245.1		PPB	05 Apr 2017 10:52:59	U	SMPL
CCB51 - 1	CCB51	-7	0.0110	-245.1		PPB	05 Apr 2017 10:55:06	U	SMPL