


ISM02.4

LB86708
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42155



A= 0.0000e+000

B= 2.3630e-004

C= -2.8453e-003

Rho= 0.9998505

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.026	0.026	124	0.000	124					-
0.2	0.200	0.218	0.018	935	0.0 %	935					9
2.5	2.500	2.456	-0.044	10407	0.0 %	10407					-2
5.0	5.000	4.923	-0.077	20847	0.0 %	20847					-2
7.5	7.500	7.618	0.118	32250	0.0 %	32250					2
10.0	10.000	9.958	-0.042	42155	0.0 %	42155					0

ns
4/5/17

LB86708
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	124	-		0 ISM02.4	PPB	4/5/2017 11:13	Std
	0.2 S0.2	935	-		0.2 ISM02.4	PPB	4/5/2017 11:16	Std
	2.5 S2.5	10407	-		2.5 ISM02.4	PPB	4/5/2017 11:18	Std
	5 S5	20847	-		5 ISM02.4	PPB	4/5/2017 11:20	Std
	7.5 S7.5	32250	-		7.5 ISM02.4	PPB	4/5/2017 11:22	Std
	10 S10	42155	-		10 ISM02.4	PPB	4/5/2017 11:28	Std
ICV	ICV	16475	3.8901 -		ISM02.4	PPB	4/5/2017 11:31	SMPL
ICB	ICB	9	-0.0007 -		ISM02.4	PPB	4/5/2017 11:33	SMPL
CCV052	CCV052	20828	4.9187 -		ISM02.4	PPB	4/5/2017 11:35	SMPL
CCB052	CCB052	3	-0.0021 -		ISM02.4	PPB	4/5/2017 11:38	SMPL
PB98047BL	PBW003	9	-0.0007 -		ISM02.4	PPB	4/5/2017 11:40	SMPL
I2468-01	MYAAW6	998	0.233 -		ISM02.4	PPB	4/5/2017 11:42	SMPL
I2468-02	MYAAX7	22	0.0024 -		ISM02.4	PPB	4/5/2017 11:44	SMPL
I2468-03	MYAAAY9	32	0.0047 -		ISM02.4	PPB	4/5/2017 11:47	SMPL
I2468-04	MYAAZ0	35	0.0054 -		ISM02.4	PPB	4/5/2017 11:49	SMPL
I2468-05	MYAAW8	227	0.0508 -		ISM02.4	PPB	4/5/2017 11:51	SMPL
I2468-06	MYAAW8D	109	0.0229 -		ISM02.4	PPB	4/5/2017 11:54	SMPL
I2468-07	MYAAW8S	4430	1.0439 -		ISM02.4	PPB	4/5/2017 11:56	SMPL
I2468-08	MYAAW9	71	0.0139 -		ISM02.4	PPB	4/5/2017 11:58	SMPL
I2468-09	MYAAZ1	115	0.0243 -		ISM02.4	PPB	4/5/2017 12:00	SMPL
PB98048BL	PBW003	13	0.0002 -		ISM02.4	PPB	4/5/2017 12:03	SMPL
I2469-01	MYAAX2	102	0.0213 -		ISM02.4	PPB	4/5/2017 12:05	SMPL
I2469-02	MYAAZ3	32	0.0047 -		ISM02.4	PPB	4/5/2017 12:07	SMPL
I2469-03	MYAB06	26	0.0033 -		ISM02.4	PPB	4/5/2017 12:09	SMPL
I2469-04	MYAB07	31	0.0045 -		ISM02.4	PPB	4/5/2017 12:12	SMPL
I2469-05	MYAAX4	71	0.0139 -		ISM02.4	PPB	4/5/2017 12:14	SMPL
I2469-06	MYAAX4D	49	0.0087 -		ISM02.4	PPB	4/5/2017 12:16	SMPL
I2469-07	MYAAX4S	4321	1.0182 -		ISM02.4	PPB	4/5/2017 12:19	SMPL
I2469-08	MYAAX5	42	0.0071 -		ISM02.4	PPB	4/5/2017 12:21	SMPL
I2469-09	MYAB08	33	0.005 -		ISM02.4	PPB	4/5/2017 12:23	SMPL
CCV053	CCV053	20862	4.9268 -		ISM02.4	PPB	4/5/2017 12:25	SMPL
CCB053	CCB053	51	0.0092 -		ISM02.4	PPB	4/5/2017 12:28	SMPL
PB98049BL	PBW003	4	-0.0019 -		ISM02.4	PPB	4/5/2017 12:30	SMPL
I2514-01	MYAAW4	90	0.0184 -		ISM02.4	PPB	4/5/2017 12:32	SMPL
I2514-02	MYAAW5	97	0.0201 -		ISM02.4	PPB	4/5/2017 12:34	SMPL
I2514-03	MYAAW7	43	0.0073 -		ISM02.4	PPB	4/5/2017 12:37	SMPL
I2514-04	MYAAX8	104	0.0217 -		ISM02.4	PPB	4/5/2017 12:39	SMPL
I2514-05	MYAAX6	47	0.0083 -		ISM02.4	PPB	4/5/2017 12:41	SMPL
I2514-06	MYAAX9	68	0.0132 -		ISM02.4	PPB	4/5/2017 12:44	SMPL
I2514-07	MYAAAY0	79	0.0158 -		ISM02.4	PPB	4/5/2017 12:46	SMPL
I2514-08	MYAB11	75	0.0149 -		ISM02.4	PPB	4/5/2017 12:48	SMPL
I2514-09	MYAAAY1	39	0.0064 -		ISM02.4	PPB	4/5/2017 12:50	SMPL
I2514-10	MYAAAY2	28	0.0038 -		ISM02.4	PPB	4/5/2017 12:53	SMPL
I2514-11	MYAAAY3	4	-0.0019 -		ISM02.4	PPB	4/5/2017 12:55	SMPL
I2514-12	MYAAAY4	166	0.0364 -		ISM02.4	PPB	4/5/2017 12:57	SMPL
I2514-13	MYAAAY5	16	0.0009 -		ISM02.4	PPB	4/5/2017 13:00	SMPL

LB86708
INSTRUMENT ID : CV1

I2514-14	MYAAY5D	19	0.0016 -	ISM02.4	PPB	4/5/2017 13:02 SMPL
I2514-15	MYAAY5S	4343	1.0234 -	ISM02.4	PPB	4/5/2017 13:04 SMPL
I2514-16	MYAB13	14	0.0005 -	ISM02.4	PPB	4/5/2017 13:06 SMPL
I2514-17	MYAB80	1140	0.2665 -	ISM02.4	PPB	4/5/2017 13:09 SMPL
I2514-18	MYAB82	237	0.0532 -	ISM02.4	PPB	4/5/2017 13:11 SMPL
I2514-19	MYAAY6	54	0.0099 -	ISM02.4	PPB	4/5/2017 13:13 SMPL
CCV054	CCV054	20890	4.9334 -	ISM02.4	PPB	4/5/2017 13:15 SMPL
CCB054	CCB054	70	0.0137 -	ISM02.4	PPB	4/5/2017 13:18 SMPL
I2514-20	MYAAY7	50	0.009 -	ISM02.4	PPB	4/5/2017 13:20 SMPL
I2514-21	MYAAY8	57	0.0106 -	ISM02.4	PPB	4/5/2017 13:22 SMPL
I2514-22	MYAB09	19	0.0016 -	ISM02.4	PPB	4/5/2017 13:25 SMPL
PB98050BL	PBW003	108	0.0227 -	ISM02.4	PPB	4/5/2017 13:27 SMPL
I2515-01	MYAAX0	156	0.034 -	ISM02.4	PPB	4/5/2017 13:29 SMPL
I2515-02	MYAAX1	174	0.0383 -	ISM02.4	PPB	4/5/2017 13:31 SMPL
I2515-03	MYAAX3	133	0.0286 -	ISM02.4	PPB	4/5/2017 13:34 SMPL
I2515-04	MYAAZ4	150	0.0326 -	ISM02.4	PPB	4/5/2017 13:36 SMPL
I2515-05	MYAAZ2	129	0.0276 -	ISM02.4	PPB	4/5/2017 13:38 SMPL
I2515-06	MYAAZ5	146	0.0317 -	ISM02.4	PPB	4/5/2017 13:41 SMPL
I2515-07	MYAAZ6	114	0.0241 -	ISM02.4	PPB	4/5/2017 13:43 SMPL
I2515-08	MYAB10	159	0.0347 -	ISM02.4	PPB	4/5/2017 13:45 SMPL
I2515-09	MYAAZ7	127	0.0272 -	ISM02.4	PPB	4/5/2017 13:47 SMPL
I2515-10	MYAAZ8	115	0.0243 -	ISM02.4	PPB	4/5/2017 13:50 SMPL
I2515-11	MYAAZ9	122	0.026 -	ISM02.4	PPB	4/5/2017 13:52 SMPL
I2515-12	MYAB00	109	0.0229 -	ISM02.4	PPB	4/5/2017 13:54 SMPL
I2515-13	MYAB81	327	0.0744 -	ISM02.4	PPB	4/5/2017 13:57 SMPL
I2515-14	MYAB02	120	0.0255 -	ISM02.4	PPB	4/5/2017 13:59 SMPL
I2515-15	MYAB02D	121	0.0257 -	ISM02.4	PPB	4/5/2017 14:01 SMPL
I2515-16	MYAB02S	4341	1.0229 -	ISM02.4	PPB	4/5/2017 14:03 SMPL
CCV055	CCV055	21051	4.9714 -	ISM02.4	PPB	4/5/2017 14:06 SMPL
CCB055	CCB055	65	0.0125 -	ISM02.4	PPB	4/5/2017 14:08 SMPL
I2515-17	MYAB12	112	0.0236 -	ISM02.4	PPB	4/5/2017 14:10 SMPL
I2515-18	MYAB79	115	0.0243 -	ISM02.4	PPB	4/5/2017 14:12 SMPL
I2515-19	MYAB01	76	0.0151 -	ISM02.4	PPB	4/5/2017 14:15 SMPL
I2515-20	MYAB03	146	0.0317 -	ISM02.4	PPB	4/5/2017 14:17 SMPL
I2515-21	MYAB04	140	0.0302 -	ISM02.4	PPB	4/5/2017 14:19 SMPL
I2515-22	MYAB05	146	0.0317 -	ISM02.4	PPB	4/5/2017 14:22 SMPL
CCV056	CCV056	21022	4.9646 -	ISM02.4	PPB	4/5/2017 14:24 SMPL
CCB056	CCB056	55	0.0102 -	ISM02.4	PPB	4/5/2017 14:26 SMPL