

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB86709

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP54006
calibration std. phosphate 0.5 ppm	WP54005
calibration std. phosphate 0.3 ppm	WP54004
calibration std. 0 ppm	WP54001
calibration std. phosphate 0.1 ppm	WP54003
calibration std. phosphate 0.05 ppm	WP54002
Combined reagent	WP54008
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate ICV std.	WP54010
phosphate CCV std.	WP54009

Intercept: -0.0007Slope: 0.6468Regression: 0.99984

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume	Final Volume	Reading	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/05/2017	13:15
2	CAL2	0.05	1	50	50	0.032	04/05/2017	13:15
3	CAL3	0.10	1	50	50	0.065	04/05/2017	13:16
4	CAL4	0.30	1	50	50	0.186	04/05/2017	13:16
5	CAL5	0.50	1	50	50	0.329	04/05/2017	13:17
6	CAL6	1.00	1	50	50	0.645	04/05/2017	13:17



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB86709

Seq	Lab ID	True Value (mg/l)	DF	Initial Vol/wt (ml/g)	Final Volume	Reading	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.497	04/05/2017	13:18
2	ICB		1	50	50	0.000	0.001	04/05/2017	13:18
3	CCV1	0.50	1	50	50	0.328	0.508	04/05/2017	13:19
4	CCB1		1	50	50	0.000	0.001	04/05/2017	13:19
5	PB98060BL		1	50	50	0.000	0.001	04/05/2017	13:20
6	PB98060BS	0.50	1	50	50	0.320	0.496	04/05/2017	13:20
7	I2527-01		1	50	50	0.063	0.098	04/05/2017	13:21
8	I2527-01DUP		1	50	50	0.062	0.097	04/05/2017	13:21
9	I2527-01MS	0.50	1	50	50	0.376	0.582	04/05/2017	13:22
10	I2527-01MSD	0.50	1	50	50	0.374	0.579	04/05/2017	13:22
11	CCV2	0.50	1	50	50	0.330	0.511	04/05/2017	13:23
12	CCB2		1	50	50	0.000	0.001	04/05/2017	13:23