


7471A

LB 86 718
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42437



A= 0.0000e+000

B= 2.3646e-004

C= -4.6723e-002

Rho= 0.9992209

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	104	0.000	104				
0.2	0.200	0.169	-0.031	912	0.0 %	912				
2.5	2.500	2.448	-0.052	10551	0.0 %	10551				
5.0	5.000	5.300	0.300	22610	0.0 %	22610				
7.5	7.500	7.318	-0.182	31145	0.0 %	31145				
10.0	10.000	9.988	-0.012	42437	0.0 %	42437				

LB86718 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 05 Apr 2017 15:50:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	104	-	0.0000	7471A	PPB	05 Apr 2017 16:07:58	S	Std
0.2 - 1	50.2	912	-	0.2000	7471A	PPB	05 Apr 2017 16:10:15	S	Std
2.5 - 1	52.5	10551	-	2.5000	7471A	PPB	05 Apr 2017 16:12:33	S	Std
5.0 - 1	55	22610	-	5.0000	7471A	PPB	05 Apr 2017 16:14:50	S	Std
7.5 - 1	57.5	31145	-	7.5000	7471A	PPB	05 Apr 2017 16:17:08	S	Std
10.0 - 1	510	42437	-	10.0000	7471A	PPB	05 Apr 2017 16:22:12	S	Std
ICV06 - 1	ICV06	17264	4.0355	-	7471A	PPB	05 Apr 2017 16:25:08	U	SMPL
ICB06 - 1	ICB06	-8	-0.0486	-	7471A	PPB	05 Apr 2017 16:27:24	U	SMPL
CCV57 - 1	CCV57	21125	4.9484	-	7471A	PPB	05 Apr 2017 16:29:39	U	SMPL
CCB57 - 1	CCB57	-7	-0.0484	-	7471A	PPB	05 Apr 2017 16:31:55	U	SMPL
CRA - 1	CRA	878	0.1609	-	7471A	PPB	05 Apr 2017 16:34:11	U	SMPL
HighStd - 1	HighStd	42723	10.0554	-	7471A	PPB	05 Apr 2017 16:36:26	U	SMPL
ChkStd - 1	ChkStd	29387	6.9020	-	7471A	PPB	05 Apr 2017 16:38:42	U	SMPL
PB98051BL - 1	PBS	2	-0.0463	-	7471A	PPB	05 Apr 2017 16:41:00	U	SMPL
PB98051BS - 1	LCSS	16603	3.8792	-	7471A	PPB	05 Apr 2017 16:43:16	U	SMPL
I2522-01 - 1	BASING-1(0-2)	3198	0.7095	-	7471A	PPB	05 Apr 2017 16:45:33	U	SMPL
I2522-03 - 1	BASING-2(0-2)	2287	0.4941	-	7471A	PPB	05 Apr 2017 16:47:51	U	SMPL
I2522-05 - 1	BASING-3(0-2)	1399	0.2841	-	7471A	PPB	05 Apr 2017 16:50:08	U	SMPL
I2522-07 - 1	BASING-4(0-2)	1064	0.2049	-	7471A	PPB	05 Apr 2017 16:52:26	U	SMPL
I2522-09 - 1	BASING-5(0-2)	1545	0.3186	-	7471A	PPB	05 Apr 2017 16:54:42	U	SMPL
CCV58 - 1	CCV58	21503	5.0378	-	7471A	PPB	05 Apr 2017 16:56:58	U	SMPL
CCB58 - 1	CCB58	39	-0.0375	-	7471A	PPB	05 Apr 2017 16:59:14	U	SMPL
I2522-11 - 1	BASING-6(0-2)	2056	0.4394	-	7471A	PPB	05 Apr 2017 17:01:31	U	SMPL
I2522-13 - 1	BASING-7(0-2)	1798	0.3784	-	7471A	PPB	05 Apr 2017 17:03:47	U	SMPL
I2522-15 - 1	BASING-8(0-2)	7857	1.8111	-	7471A	PPB	05 Apr 2017 17:06:03	U	SMPL
I2522-17 - 1	BASING-9(0-2)	1025	0.1956	-	7471A	PPB	05 Apr 2017 17:08:19	U	SMPL
I2522-19 - 1	BASING-10(0-2)	1252	0.2493	-	7471A	PPB	05 Apr 2017 17:10:36	U	SMPL
I2523-01 - 1	BASING-11(0-2)	1250	0.2488	-	7471A	PPB	05 Apr 2017 17:12:53	U	SMPL
I2523-03 - 1	BASING-12(0-2)	1000	0.1897	-	7471A	PPB	05 Apr 2017 17:15:10	U	SMPL
I2523-05 - 1	BASING-13(0-2)	986	0.1864	-	7471A	PPB	05 Apr 2017 17:17:28	U	SMPL
I2523-07 - 1	BASING-14(0-2)	530	0.0786	-	7471A	PPB	05 Apr 2017 17:19:46	U	SMPL
I2523-09 - 1	BASING-15(0-2)	907	0.1677	-	7471A	PPB	05 Apr 2017 17:22:02	U	SMPL
CCV59 - 1	CCV59	21418	5.0177	-	7471A	PPB	05 Apr 2017 17:24:19	U	SMPL
CCB59 - 1	CCB59	56	-0.0335	-	7471A	PPB	05 Apr 2017 17:26:35	U	SMPL
I2523-11 - 1	CEMETERY-1(0-7)	773	0.1361	-	7471A	PPB	05 Apr 2017 17:28:51	U	SMPL
I2523-13 - 1	CEMETERY-2(0-7)	960	0.1803	-	7471A	PPB	05 Apr 2017 17:33:19	U	SMPL
I2523-15 - 1	CEMETERY-3(0-7)	370	0.0408	-	7471A	PPB	05 Apr 2017 17:35:34	U	SMPL
I2523-15DUP - 1	CEMETERY-3(0-7)DUP	353	0.0367	-	7471A	PPB	05 Apr 2017 17:37:50	U	SMPL
I2523-15MS - 1	CEMETERY-3(0-7)MS	15844	3.6997	-	7471A	PPB	05 Apr 2017 17:40:07	U	SMPL
I2523-15MSD - 1	CEMETERY-3(0-7)MSD	15937	3.7217	-	7471A	PPB	05 Apr 2017 17:42:24	U	SMPL
I2523-15LX5 - 1		126	-0.0169	-	7471A	PPB	05 Apr 2017 17:44:41	U	SMPL
I2523-15A - 1		17525	4.0972	-	7471A	PPB	05 Apr 2017 17:46:59	U	SMPL
CCV60 - 1	CCV60	21556	5.0503	-	7471A	PPB	05 Apr 2017 17:49:17	U	SMPL
CCB60 - 1	CCB60	38	-0.0377	-	7471A	PPB	05 Apr 2017 17:51:33	U	SMPL