



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86722

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/05/2017	16:25
2	CAL2	1	Water			7.00	04/05/2017	16:26
3	CAL3	1	Water			10.02	04/05/2017	16:27
4	ICV	1	Water			7.00	04/05/2017	16:28
5	CCV1	1	Water			2.02	04/05/2017	16:30
6	I2569-03	1	Solid	20.03	20	7.50	04/05/2017	16:31
7	I2569-03DUP	1	Solid	20.05	20	7.51	04/05/2017	16:32
8	I2571-03	1	Solid	20.02	20	8.45	04/05/2017	16:33
9	I2572-03	1	Solid	20.04	20	7.10	04/05/2017	16:34
10	I2572-06	1	Solid	20.01	20	7.12	04/05/2017	16:35
11	I2572-09	1	Solid	20.05	20	6.51	04/05/2017	16:36
12	CCV2	1	Water			12.01	04/05/2017	16:37

1386722

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I2569

WorkList ID : 97161

Date : 4/5/2017 4:08:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2017	Solid	I2569-03	Corrosivity	Cool 4 deg C	PSEG01	N51	RT3456-275030	04/05/2017	9045C
04/10/2017	Solid	I2571-03	Corrosivity	Cool 4 deg C	PSEG01	N51	DUNCAN-AVE	04/05/2017	9045C
04/10/2017	Solid	I2572-03	Corrosivity	Cool 4 deg C	PSEG01	N51	15-41	04/05/2017	9045C
04/10/2017	Solid	I2572-06	Corrosivity	Cool 4 deg C	PSEG01	N51	RT-2891	04/05/2017	9045C
04/10/2017	Solid	I2572-09	Corrosivity	Cool 4 deg C	PSEG01	N51	RT-3460	04/05/2017	9045C

Date/Time 04/05/17 4:08 PM
Received by: JD
Relinquished by: JD

Date/Time 04/05/17 6:46 PM
Received by: CP
Relinquished by: SD