


7470A

LB 86728
INSTRUMENT ID: CV1

Linear

 μ Abs.:

3464E



A= 0.0000e+000

B= 2.8495e-004

C= 1.7183e-002

Rho= 0.9997126

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	26	0.000	26				
0.2	0.200	0.196	-0.004	627	0.0 %	627				
2.5	2.500	2.425	-0.075	8451	0.0 %	8451				
5.0	5.000	4.996	-0.004	17473	0.0 %	17473				
7.5	7.500	7.669	0.169	26853	0.0 %	26853				
10.0	10.000	9.889	-0.111	34646	0.0 %	34646				

LB86728 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Apr 2017 11:00:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	06 Apr 2017 11:04:06	S	Std
0.2 - 1	50.2	627	-	0.2000	7470A	PPB	06 Apr 2017 11:06:15	S	Std
2.5 - 1	52.5	8451	-	2.5000	7470A	PPB	06 Apr 2017 11:08:23	S	Std
5.0 - 1	55	17473	-	5.0000	7470A	PPB	06 Apr 2017 11:10:31	S	Std
7.5 - 1	57.5	26853	-	7.5000	7470A	PPB	06 Apr 2017 11:15:08	S	Std
10.0 - 1	510	34646	-	10.0000	7470A	PPB	06 Apr 2017 11:17:15	S	Std
ICV07 - 1	ICV07	14100	4.0349	-	7470A	PPB	06 Apr 2017 11:22:23	U	SMPL
ICB07 - 1	ICB07	-84	-0.0068	-	7470A	PPB	06 Apr 2017 11:24:29	U	SMPL
CCV61 - 1	CCV61	17785	5.0849	-	7470A	PPB	06 Apr 2017 11:29:05	U	SMPL
CCB61 - 1	CCB61	-104	-0.0125	-	7470A	PPB	06 Apr 2017 11:31:11	U	SMPL
CRA - 1	CRA	626	0.1956	-	7470A	PPB	06 Apr 2017 11:33:18	U	SMPL
HighStd - 1	HighStd	35261	10.0647	-	7470A	PPB	06 Apr 2017 11:35:24	U	SMPL
ChkStd - 1	ChkStd	25253	7.2129	-	7470A	PPB	06 Apr 2017 11:37:31	U	SMPL
PB98071BL - 1	PBW	-143	-0.0236	-	7470A	PPB	06 Apr 2017 11:39:40	U	SMPL
PB98071BS - 1	LCSW	13978	4.0002	-	7470A	PPB	06 Apr 2017 11:41:48	U	SMPL
PB98071TB - 1	PB98052TB	-72	-0.0033	-	7470A	PPB	06 Apr 2017 11:43:56	U	SMPL
I2537-05 - 1	B1-(0-5)	5	0.0186	-	7470A	PPB	06 Apr 2017 11:46:04	U	SMPL
I2537-10 - 1	B4-(0-5)	14	0.0212	-	7470A	PPB	06 Apr 2017 11:48:12	U	SMPL
I2537-15 - 1	B5-(0-5)	86	0.0417	-	7470A	PPB	06 Apr 2017 11:50:21	U	SMPL
I2537-20 - 1	B8-(0-5)	101	0.0460	-	7470A	PPB	06 Apr 2017 11:52:27	U	SMPL
CCV62 - 1	CCV62	17504	5.0049	-	7470A	PPB	06 Apr 2017 11:54:34	U	SMPL
CCB62 - 1	CCB62	-94	-0.0096	-	7470A	PPB	06 Apr 2017 11:56:41	U	SMPL
I2538-04 - 1	LOCUST-ST-CONCRETE	113	0.0494	-	7470A	PPB	06 Apr 2017 11:58:48	U	SMPL
I2538-10 - 1	LOCUST-ST-IP(3-5)	144	0.0582	-	7470A	PPB	06 Apr 2017 12:00:55	U	SMPL
I2538-10DUP - 1	LOCUST-ST-IP(3-5)DUP	132	0.0548	-	7470A	PPB	06 Apr 2017 12:03:02	U	SMPL
I2538-10MS - 1	LOCUST-ST-IP(3-5)MS	12550	3.5933	-	7470A	PPB	06 Apr 2017 12:05:09	U	SMPL
I2538-10MSD - 1	LOCUST-ST-IP(3-5)MSD	13059	3.7383	-	7470A	PPB	06 Apr 2017 12:07:16	U	SMPL
PB98072BL - 1	PBW	-97	-0.0105	-	7470A	PPB	06 Apr 2017 12:09:24	U	SMPL
PB98072BS - 1	LCSW	13314	3.8110	-	7470A	PPB	06 Apr 2017 12:11:32	U	SMPL
I2532-04 - 1	49B	1602	0.4737	-	7470A	PPB	06 Apr 2017 12:18:37	U	SMPL
I2532-04DUP - 1	49BDUP	1544	0.4571	-	7470A	PPB	06 Apr 2017 12:20:44	U	SMPL
I2532-04MS - 1	49BMS	16456	4.7063	-	7470A	PPB	06 Apr 2017 12:22:51	U	SMPL
CCV63 - 1	CCV63	18832	5.3833	-	7470A	PPB	06 Apr 2017 12:24:58	U	SMPL
CCB63 - 1	CCB63	-23	0.0106	-	7470A	PPB	06 Apr 2017 12:27:05	U	SMPL
I2532-04MSD - 1	49BMSD	17034	4.8710	-	7470A	PPB	06 Apr 2017 12:29:12	U	SMPL
I2532-05 - 1	53A	652	0.2030	-	7470A	PPB	06 Apr 2017 12:31:19	U	SMPL
I2532-06 - 1	54C	135	0.0557	-	7470A	PPB	06 Apr 2017 12:33:26	U	SMPL
I2532-07 - 1	37A	631	0.1970	-	7470A	PPB	06 Apr 2017 12:35:33	U	SMPL
I2532-08 - 1	38D	81	0.0403	-	7470A	PPB	06 Apr 2017 12:37:40	U	SMPL
I2532-09 - 1	FIELD DUPLICATE	794	0.2434	-	7470A	PPB	06 Apr 2017 12:39:47	U	SMPL
I2532-10 - 1	FIELD BLANK	59	0.0340	-	7470A	PPB	06 Apr 2017 12:41:56	U	SMPL
CCV64 - 1	CCV64	17857	5.1055	-	7470A	PPB	06 Apr 2017 12:44:04	U	SMPL
CCB64 - 1	CCB64	-33	0.0078	-	7470A	PPB	06 Apr 2017 12:46:12	U	SMPL
I2538-10LX5 - 1		34	0.0269	-	7470A	PPB	06 Apr 2017 12:48:19	U	SMPL
I2538-10A - 1		15029	4.2996	-	7470A	PPB	06 Apr 2017 12:50:26	U	SMPL
I2532-04LX5 - 1		483	0.1548	-	7470A	PPB	06 Apr 2017 12:55:28	U	SMPL
I2532-04A - 1		17844	5.1018	-	7470A	PPB	06 Apr 2017 12:57:35	U	SMPL
CCV65 - 1	CCV65	18151	5.1892	-	7470A	PPB	06 Apr 2017 12:59:42	U	SMPL
CCB65 - 1	CCB65	-41	0.0055	-	7470A	PPB	06 Apr 2017 13:01:49	U	SMPL