

245.1


LB 86730

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34734



A= 0.0000e+000

B= 2.8862e-004

C= -9.3253e-003

Rho= 0.9998652

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	26	0.000	26				
0.2	0.200	0.190	-0.010	689	0.0 %	689				
2.5	2.500	2.588	0.088	8999	0.0 %	8999				
5.0	5.000	4.884	-0.116	16956	0.0 %	16956				
7.5	7.500	7.524	0.024	26103	0.0 %	26103				
10.0	10.000	10.015	0.015	34734	0.0 %	34734				

LB86730 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 06 Apr 2017 13:11:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	245.1	PPB	06 Apr 2017 13:56:37	S	Std
0.2 - 1	30.2	689	-	0.2000	245.1	PPB	06 Apr 2017 13:58:45	S	Std
2.5 - 1	52.5	8999	-	2.5000	245.1	PPB	06 Apr 2017 14:00:54	S	Std
5.0 - 1	55	16956	-	5.0000	245.1	PPB	06 Apr 2017 14:03:02	S	Std
7.5 - 1	57.5	26103	-	7.5000	245.1	PPB	06 Apr 2017 14:05:10	S	Std
10.0 - 1	510	34734	-	10.0000	245.1	PPB	06 Apr 2017 14:07:17	S	Std
ICV08 - 1	ICV08	13565	3.9058	-245.1	PPB	PPB	06 Apr 2017 14:10:02	U	SMPL
ICB08 - 1	ICB08	-52	-0.0243	-245.1	PPB	PPB	06 Apr 2017 14:12:08	U	SMPL
CCV66 - 1	CCV66	16694	4.8088	-245.1	PPB	PPB	06 Apr 2017 14:14:15	U	SMPL
CCB66 - 1	CCB66	-60	-0.0266	-245.1	PPB	PPB	06 Apr 2017 14:16:21	U	SMPL
CRA - 1	CRA	676	0.1858	-245.1	PPB	PPB	06 Apr 2017 14:18:28	U	SMPL
HighStd - 1	HighStd	34603	9.9777	-245.1	PPB	PPB	06 Apr 2017 14:20:35	U	SMPL
ChkStd - 1	ChkStd	24148	6.9602	-245.1	PPB	PPB	06 Apr 2017 14:22:41	U	SMPL
PB98070BL - 1	PBW	-59	-0.0264	-245.1	PPB	PPB	06 Apr 2017 14:24:49	U	SMPL
PB98070BS - 1	LCSW	13429	3.8665	-245.1	PPB	PPB	06 Apr 2017 14:26:57	U	SMPL
I2529-01 - 1	S1-JOINT3	48	0.0045	-245.1	PPB	PPB	06 Apr 2017 14:29:05	U	SMPL
I2529-01DUP - 1	S1-JOINT3DUP	64	0.0091	-245.1	PPB	PPB	06 Apr 2017 14:31:13	U	SMPL
I2529-01MS - 1	S1-JOINT3MS	13317	3.8342	-245.1	PPB	PPB	06 Apr 2017 14:33:21	U	SMPL
I2529-01MSD - 1	S1-JOINT3MSD	13858	3.9903	-245.1	PPB	PPB	06 Apr 2017 14:35:31	U	SMPL
I2529-02 - 1	S2-MIDSTREAM	17	-0.0044	-245.1	PPB	PPB	06 Apr 2017 14:37:38	U	SMPL
CCV67 - 1	CCV67	16606	4.7834	-245.1	PPB	PPB	06 Apr 2017 14:39:45	U	SMPL
CCB67 - 1	CCB67	-50	-0.0238	-245.1	PPB	PPB	06 Apr 2017 14:41:52	U	SMPL
I2529-03 - 1	S3-DOWNSTREAM	9	-0.0067	-245.1	PPB	PPB	06 Apr 2017 14:43:58	U	SMPL
I2529-04 - 1	S4-UPSTREAM	17	-0.0044	-245.1	PPB	PPB	06 Apr 2017 14:46:05	U	SMPL
I2529-01LX5 - 1		32	-0.0001	-245.1	PPB	PPB	06 Apr 2017 14:48:12	U	SMPL
I2529-01A - 1		15100	4.3488	-245.1	PPB	PPB	06 Apr 2017 14:50:18	U	SMPL
CCV68 - 1	CCV68	16538	4.7638	-245.1	PPB	PPB	06 Apr 2017 14:52:26	U	SMPL
CCB68 - 1	CCB68	-63	-0.0275	-245.1	PPB	PPB	06 Apr 2017 14:54:34	U	SMPL