



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB86733

Temp.: 21.3 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/06/2017	14:00
2	CAL2	1	Water			7.02	04/06/2017	14:01
3	CAL3	1	Water			10.03	04/06/2017	14:02
4	ICV	1	Water			7.01	04/06/2017	14:03
5	CCV1	1	Water			2.02	04/06/2017	14:05
6	I2582-01	1	Water			6.76	04/06/2017	14:06
7	I2582-01DUP	1	Water			6.77	04/06/2017	14:07
8	I2582-04	1	Water			7.42	04/06/2017	14:08
9	CCV2	1	Water			12.01	04/06/2017	14:10

13 86733

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I2582

WorkList ID : 97198

Date : 4/6/2017 1:38:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/06/2017	Water	I2582-01-6	pH	Cool 4 deg C	PSEG01	O11	GW-1	04/06/2017	9045C
04/06/2017	Water	I2582-04 -6	pH	Cool 4 deg C	PSEG01	O11	GW-2	04/06/2017	9045C

Date/Time 64106117113809
Received by: SP
Relinquished by: SP

Date/Time 64106117 310000
Received by: SP
Relinquished by: SP