


7471A

LB 86739
INSTRUMENT ID: CV4

Linear

 μ Abs.:

41514



A= 0.0000e+000

B= 2.3812e-004

C= -1.3710e-002

Rho= 0.9997182

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	16	0.000	16				
0.2	0.200	0.186	-0.014	839	0.0 %	839				
2.5	2.500	2.477	-0.023	10462	0.0 %	10462				
5.0	5.000	5.005	0.005	21078	0.0 %	21078				
7.5	7.500	7.669	0.169	32266	0.0 %	32266				
10.0	10.000	9.872	-0.128	41514	0.0 %	41514				

LB86739 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 06 Apr 2017 15:02:19					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	16	-	0.00007471A	PPB		06 Apr 2017 15:04:26	S	Std
0.2 - 1	S0.2	839	-	0.20007471A	PPB		06 Apr 2017 15:06:44	S	Std
2.5 - 1	S2.5	10462	-	2.50007471A	PPB		06 Apr 2017 15:09:02	S	Std
5.0 - 1	S5	21078	-	5.00007471A	PPB		06 Apr 2017 15:11:19	S	Std
7.5 - 1	S7.5	32266	-	7.50007471A	PPB		06 Apr 2017 15:13:36	S	Std
10.0 - 1	S10	41514	-	10.00007471A	PPB		06 Apr 2017 15:15:53	S	Std
ICV09 - 1	ICV09	17848	4.2362	-7471A	PPB		06 Apr 2017 15:18:38	U	SMPL
ICB09 - 1	ICB09	-68	-0.0299	-7471A	PPB		06 Apr 2017 15:20:53	U	SMPL
CCV69 - 1	CCV69	21729	5.1604	-7471A	PPB		06 Apr 2017 15:23:09	U	SMPL
CCB69 - 1	CCB69	-77	-0.0320	-7471A	PPB		06 Apr 2017 15:25:24	U	SMPL
CRA - 1	CRA	826	0.1830	-7471A	PPB		06 Apr 2017 15:27:40	U	SMPL
HighStd - 1	HighStd	41130	9.7801	-7471A	PPB		06 Apr 2017 15:29:56	U	SMPL
ChkStd - 1	ChkStd	29967	7.1220	-7471A	PPB		06 Apr 2017 15:32:12	U	SMPL
PB98081BL - 1	PBS	-103	-0.0382	-7471A	PPB		06 Apr 2017 15:34:29	U	SMPL
PB98081BS - 1	LCSS	14848	3.5219	-7471A	PPB		06 Apr 2017 15:36:46	U	SMPL
I2532-02 - 1	48A-9-11	1309	0.2980	-7471A	PPB		06 Apr 2017 15:39:03	U	SMPL
I2533-01 - 1	CT-MIXCOMP-1	34318	8.1580	-7471A	PPB		06 Apr 2017 15:41:21	U	SMPL
I2536-01 - 1	E2-BB7-BASE	88745	21.1181	-7471A	PPB		06 Apr 2017 15:43:38	U	SMPL
I2536-02 - 1	E2-BB8-BASE	617690	HIGH	-7471A	PPB		06 Apr 2017 15:45:56	X	
I2537-01 - 1	B1-(0-5)	14422	3.4204	-7471A	PPB		06 Apr 2017 15:47:47	U	SMPL
CCV70 - 1	CCV70	21802	5.1778	-7471A	PPB		06 Apr 2017 15:50:06	U	SMPL
CCB70 - 1	CCB70	-61	-0.0282	-7471A	PPB		06 Apr 2017 15:52:22	U	SMPL
I2537-01DUP - 1	B1-(0-5)DUP	14859	3.5245	-7471A	PPB		06 Apr 2017 15:54:38	U	SMPL
I2537-01MS - 1	B1-(0-5)MS	34119	8.1107	-7471A	PPB		06 Apr 2017 15:56:55	U	SMPL
I2537-01MSD - 1	B1-(0-5)MSD	35010	8.3228	-7471A	PPB		06 Apr 2017 15:59:11	U	SMPL
I2537-06 - 1	B4-(0-5)	2899	0.6766	-7471A	PPB		06 Apr 2017 16:01:27	U	SMPL
I2537-11 - 1	B5-(0-5)	5420	1.2769	-7471A	PPB		06 Apr 2017 16:03:43	U	SMPL
I2537-16 - 1	B8-(0-5)	18606	4.4167	-7471A	PPB		06 Apr 2017 16:06:00	U	SMPL
I2538-01 - 1	LOCUST-ST-CONCRETE	244	0.0444	-7471A	PPB		06 Apr 2017 16:08:18	U	SMPL
I2538-06 - 1	LOCUST-ST-IP(3-5)	36436	8.6624	-7471A	PPB		06 Apr 2017 16:10:35	U	SMPL
I2542-14 - 1	COMP-1	1706	0.3925	-7471A	PPB		06 Apr 2017 16:12:53	U	SMPL
I2542-15 - 1	COMP-2	5079	1.1957	-7471A	PPB		06 Apr 2017 16:15:09	U	SMPL
CCV71 - 1	CCV71	21982	5.2206	-7471A	PPB		06 Apr 2017 16:17:26	U	SMPL
CCB71 - 1	CCB71	-16	-0.0175	-7471A	PPB		06 Apr 2017 16:19:42	U	SMPL
I2542-16 - 1	COMP-3	3590	0.8411	-7471A	PPB		06 Apr 2017 16:21:58	U	SMPL
I2542-17 - 1	COMP-4	3594	0.8421	-7471A	PPB		06 Apr 2017 16:24:15	U	SMPL
I2542-18 - 1	COMP-5	3041	0.7104	-7471A	PPB		06 Apr 2017 16:26:31	U	SMPL
I2563-01 - 1	E2-AA10-BASE	30702	7.2970	-7471A	PPB		06 Apr 2017 16:28:47	U	SMPL
I2563-02 - 1	E2-AA10-NSW	51111	12.1568	-7471A	PPB		06 Apr 2017 16:31:04	U	SMPL
I2563-03 - 1	E2-AA10-SSW	31647	7.5220	-7471A	PPB		06 Apr 2017 16:33:21	U	SMPL
I2563-04 - 1	E2-AA10-ESW	4104	0.9635	-7471A	PPB		06 Apr 2017 16:35:38	U	SMPL
PB98082BL - 1	PBS	4	-0.0128	-7471A	PPB		06 Apr 2017 16:37:56	U	SMPL
PB98082BS - 1	LCSS	15720	3.7295	-7471A	PPB		06 Apr 2017 16:40:13	U	SMPL
I2528-02 - 1	TS-3	2682	0.6249	-7471A	PPB		06 Apr 2017 16:42:30	U	SMPL
CCV72 - 1	CCV72	21454	5.0949	-7471A	PPB		06 Apr 2017 16:44:46	U	SMPL
CCB72 - 1	CCB72	-44	-0.0242	-7471A	PPB		06 Apr 2017 16:47:02	U	SMPL
I2528-04 - 1	TS-5	1632	0.3749	-7471A	PPB		06 Apr 2017 16:49:18	U	SMPL
I2528-05 - 1	TS-6	8304	1.9636	-7471A	PPB		06 Apr 2017 16:51:35	U	SMPL
I2534-01 - 1	GP-B-01(0-5)	1514	0.3468	-7471A	PPB		06 Apr 2017 16:53:51	U	SMPL
I2534-02 - 1	GP-B-01(0-5)D	1129	0.2551	-7471A	PPB		06 Apr 2017 16:56:08	U	SMPL
I2534-03 - 1	GP-B-01(5-10)	280	0.0530	-7471A	PPB		06 Apr 2017 16:58:25	U	SMPL
I2534-04 - 1	GP-B-01(10-15)	394	0.0801	-7471A	PPB		06 Apr 2017 17:00:41	U	SMPL
I2534-04DUP - 1	GP-B-01(10-15)DUP	414	0.0849	-7471A	PPB		06 Apr 2017 17:02:59	U	SMPL
I2534-04MS - 1	GP-B-01(10-15)MS	16361	3.8822	-7471A	PPB		06 Apr 2017 17:05:16	U	SMPL
I2534-04MSD - 1	GP-B-01(10-15)MSD	16417	3.8955	-7471A	PPB		06 Apr 2017 17:07:34	U	SMPL
I2534-07 - 1	GP-B-03(0-5)	563	0.1204	-7471A	PPB		06 Apr 2017 17:09:50	U	SMPL
CCV73 - 1	CCV73	21420	5.0868	-7471A	PPB		06 Apr 2017 17:12:07	U	SMPL
CCB73 - 1	CCB73	-21	-0.0187	-7471A	PPB		06 Apr 2017 17:14:23	U	SMPL
I2534-08 - 1	GP-B-03(5-10)	1146	0.2592	-7471A	PPB		06 Apr 2017 17:16:39	U	SMPL
I2534-10 - 1	GP-B-03(10-15)	1033	0.2323	-7471A	PPB		06 Apr 2017 17:18:56	U	SMPL
I2534-11 - 1	GP-B-02(0-5)	690	0.1506	-7471A	PPB		06 Apr 2017 17:21:12	U	SMPL
I2534-12 - 1	GP-B-02(5-10)	725	0.1589	-7471A	PPB		06 Apr 2017 17:23:29	U	SMPL
I2534-13 - 1	GP-B-02(10-15)	505	0.1065	-7471A	PPB		06 Apr 2017 17:25:45	U	SMPL
I2569-01 - 1	RT3456-275030	2937	0.6856	-7471A	PPB		06 Apr 2017 17:28:02	U	SMPL
I2571-01 - 1	DUNCAN-AVE	39833	9.4713	-7471A	PPB		06 Apr 2017 17:30:19	U	SMPL
I2572-01 - 1	15-41	1486	0.3401	-7471A	PPB		06 Apr 2017 17:32:37	U	SMPL
I2572-04 - 1	RT-2891	315	0.0613	-7471A	PPB		06 Apr 2017 17:34:54	U	SMPL
I2572-07 - 1	RT-3460	1301	0.2961	-7471A	PPB		06 Apr 2017 17:37:11	U	SMPL
CCV74 - 1	CCV74	21352	5.0706	-7471A	PPB		06 Apr 2017 17:39:28	U	SMPL
CCB74 - 1	CCB74	-21	-0.0187	-7471A	PPB		06 Apr 2017 17:41:45	U	SMPL
I2536-01DLX5 - 1		17528	4.1600	-7471A	PPB		06 Apr 2017 17:44:02	U	SMPL
I2536-02DLX10 - 1		121560	28.9320	-7471A	PPB		06 Apr 2017 17:46:18	U	SMPL
I2563-02DLX5 - 1		9430	2.2318	-7471A	PPB		06 Apr 2017 17:48:35	U	SMPL
I2537-01LX5 - 1		2765	0.6447	-7471A	PPB		06 Apr 2017 17:50:53	U	SMPL
I2537-01A - 1		32105	7.6311	-7471A	PPB		06 Apr 2017 17:53:10	U	SMPL
I2534-04LX5 - 1		40	-0.0042	-7471A	PPB		06 Apr 2017 17:55:26	U	SMPL
I2534-04A - 1		18028	4.2791	-7471A	PPB		06 Apr 2017 17:57:44	U	SMPL
I2536-02DL2X50 - 1		24593	5.8423	-7471A	PPB		06 Apr 2017 18:00:01	U	SMPL
CCV75 - 1	CCV75	21774	5.1711	-7471A	PPB		06 Apr 2017 18:02:19	U	SMPL
CCB75 - 1	CCB75	-25	-0.0197	-7471A	PPB		06 Apr 2017 18:04:36	U	SMPL