


7470A

LB 86753
INSTRUMENT ID: CV1

Linear

 μ Abs.:

38386



A= 0.0000e+000

B= 2.6076e-004

C= 2.6617e-002

Rho= 0.9999685

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	0	0.000	0				
0.2	0.200	0.222	0.022	749	0.0 %	749				
2.5	2.500	2.458	-0.042	9326	0.0 %	9326				
5.0	5.000	4.977	-0.023	18986	0.0 %	18986				
7.5	7.500	7.480	-0.020	28582	0.0 %	28582				
10.0	10.000	10.036	0.036	38386	0.0 %	38386				

LB86753 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Apr 2017 11:24:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.1	0	-	0.0000	7470A	PPB	07 Apr 2017 11:31:49	S	Std
0.2 - 1	50.2	749	-	0.2000	7470A	PPB	07 Apr 2017 11:33:57	S	Std
2.5 - 1	52.5	9326	-	2.5000	7470A	PPB	07 Apr 2017 11:36:06	S	Std
5.0 - 1	55	18986	-	5.0000	7470A	PPB	07 Apr 2017 11:38:14	S	Std
7.5 - 1	57.5	28582	-	7.5000	7470A	PPB	07 Apr 2017 11:40:22	S	Std
10.0 - 1	510	38386	-	10.0000	7470A	PPB	07 Apr 2017 11:44:55	S	Std
ICV10 - 1	ICV10	15110	3.9667	-7470A	PPB	07 Apr 2017 11:47:26	U	SMPL	
ICB10 - 1	ICB10	-88	0.0037	-7470A	PPB	07 Apr 2017 11:49:32	U	SMPL	
CCV76 - 1	CCV76	19185	5.0293	-7470A	PPB	07 Apr 2017 11:51:38	U	SMPL	
CCB76 - 1	CCB76	-83	0.0050	-7470A	PPB	07 Apr 2017 11:53:45	U	SMPL	
CRA - 1	CRA	756	0.2237	-7470A	PPB	07 Apr 2017 11:55:53	U	SMPL	
HighStd - 1	HighStd	38888	10.1670	-7470A	PPB	07 Apr 2017 11:57:59	U	SMPL	
ChkStd - 1	ChkStd	27305	7.1466	-7470A	PPB	07 Apr 2017 12:00:06	U	SMPL	
PB98113BL - 1	PBW	-123	-0.0055	-7470A	PPB	07 Apr 2017 12:02:15	U	SMPL	
PB98113BS - 1	LCSW	14333	3.7641	-7470A	PPB	07 Apr 2017 12:04:23	U	SMPL	
PB98113TB - 1	PB98078TB	-81	0.0055	-7470A	PPB	07 Apr 2017 12:06:31	U	SMPL	
I2569-03 - 1	RT3456-275030	6	0.0282	-7470A	PPB	07 Apr 2017 12:08:38	U	SMPL	
I2571-03 - 1	DUNCAN-AVE	27	0.0337	-7470A	PPB	07 Apr 2017 12:10:46	U	SMPL	
I2572-03 - 1	15-41	10	0.0292	-7470A	PPB	07 Apr 2017 12:12:55	U	SMPL	
I2572-06 - 1	RT-2891	17	0.0310	-7470A	PPB	07 Apr 2017 12:15:02	U	SMPL	
CCV77 - 1	CCV77	19311	5.0621	-7470A	PPB	07 Apr 2017 12:17:09	U	SMPL	
CCB77 - 1	CCB77	-98	0.0011	-7470A	PPB	07 Apr 2017 12:19:16	U	SMPL	
I2572-09 - 1	RT-3460	67	0.0441	-7470A	PPB	07 Apr 2017 12:25:55	U	SMPL	
I2572-09DUP - 1	RT-3460DUP	54	0.0407	-7470A	PPB	07 Apr 2017 12:28:01	U	SMPL	
I2572-09MS - 1	RT-3460MS	14595	3.8324	-7470A	PPB	07 Apr 2017 12:30:08	U	SMPL	
I2572-09MSD - 1	RT-3460MSD	14436	3.7909	-7470A	PPB	07 Apr 2017 12:32:15	U	SMPL	
PB98114BL - 1	PBW	-44	0.0151	-7470A	PPB	07 Apr 2017 12:34:23	U	SMPL	
PB98114BS - 1	LCSW	14217	3.7338	-7470A	PPB	07 Apr 2017 12:36:30	U	SMPL	
I2582-01 - 1	GW-1	386	0.1273	-7470A	PPB	07 Apr 2017 12:43:18	U	SMPL	
I2582-01DUP - 1	GW-1DUP	364	0.1215	-7470A	PPB	07 Apr 2017 12:45:25	U	SMPL	
I2582-01MS - 1	GW-1MS	15990	4.1961	-7470A	PPB	07 Apr 2017 12:47:34	U	SMPL	
I2582-01MSD - 1	GW-1MSD	16642	4.3661	-7470A	PPB	07 Apr 2017 12:49:41	U	SMPL	
CCV78 - 1	CCV78	19213	5.0366	-7470A	PPB	07 Apr 2017 12:51:48	U	SMPL	
CCB78 - 1	CCB78	-53	0.0128	-7470A	PPB	07 Apr 2017 12:53:55	U	SMPL	
I2582-04 - 1	GW-2	112901	29.4664	-7470A	PPB	07 Apr 2017 12:56:02	U	SMPL	
I2585-01 - 1	22A	217	0.0832	-7470A	PPB	07 Apr 2017 13:00:19	U	SMPL	
I2585-02 - 1	14A	371	0.1234	-7470A	PPB	07 Apr 2017 13:02:26	U	SMPL	
I2585-03 - 1	1A	446	0.1429	-7470A	PPB	07 Apr 2017 13:04:33	U	SMPL	
I2572-09LX5 - 1		50	0.0397	-7470A	PPB	07 Apr 2017 13:06:40	U	SMPL	
I2572-09A - 1		15049	3.9508	-7470A	PPB	07 Apr 2017 13:08:48	U	SMPL	
I2582-01LX5 - 1		-34	0.0178	-7470A	PPB	07 Apr 2017 13:10:56	U	SMPL	
I2582-01A - 1		15789	4.1437	-7470A	PPB	07 Apr 2017 13:13:04	U	SMPL	
I2582-04DLX5 - 1		20271	5.3124	-7470A	PPB	07 Apr 2017 13:15:13	U	SMPL	
CCV79 - 1	CCV79	19650	5.1505	-7470A	PPB	07 Apr 2017 13:17:20	U	SMPL	
CCB79 - 1	CCB79	-54	0.0125	-7470A	PPB	07 Apr 2017 13:19:27	U	SMPL	