

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB86759

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP53540
calibration std. phosphate 1 ppm	WP54068
calibration std. phosphate 0.5 ppm	WP54067
calibration std. phosphate 0.3 ppm	WP54066
calibration std. 0 ppm	WP54062
phosphate ICV std.	WP54070
calibration std. phosphate 0.1 ppm	WP54065
calibration std. phosphate 0.05 ppm	WP54064
Combined reagent	WP54073
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
phosphate ICV std.	WP54070

Intercept: -0.0008Slope: 0.6491Regression: 0.999804

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	04/07/2017	11:15
2	CAL2	0.05	1	50	50	0.032	04/07/2017	11:15
3	CAL3	0.10	1	50	50	0.065	04/07/2017	11:16
4	CAL4	0.30	1	50	50	0.186	04/07/2017	11:16
5	CAL5	0.50	1	50	50	0.331	04/07/2017	11:17
6	CAL6	1.00	1	50	50	0.647	04/07/2017	11:17



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB86759

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.494	04/07/2017	11:18
2	ICB		1	50	50	0.000	0.001	04/07/2017	11:18
3	CCV1	0.50	1	50	50	0.327	0.505	04/07/2017	11:19
4	CCB1		1	50	50	0.000	0.001	04/07/2017	11:19
5	PB98135BL		1	50	50	0.000	0.001	04/07/2017	11:20
6	PB98135BS	0.50	1	50	50	0.319	0.493	04/07/2017	11:20
7	I2559-02		1	50	50	1.528	2.355	04/07/2017	11:21
8	I2561-12		1	50	50	0.120	0.186	04/07/2017	11:21
9	I2561-12DUP		1	50	50	0.123	0.191	04/07/2017	11:22
10	I2561-12MS	0.50	1	50	50	0.452	0.698	04/07/2017	11:22
11	I2561-12MSD	0.50	1	50	50	0.456	0.704	04/07/2017	11:23
12	I2587-01		1	50	50	1.329	2.049	04/07/2017	11:23
13	I2559-02		10	50	50	0.578	0.892	04/07/2017	11:24
14	I2587-01		10	50	50	0.443	0.684	04/07/2017	11:24
15	CCV2	0.50	1	50	50	0.329	0.508	04/07/2017	11:25
16	CCB2		1	50	50	0.000	0.001	04/07/2017	11:25