



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86760

Temp.: 19.4 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/07/2017	14:10
2	CAL2	1	Water			7.00	04/07/2017	14:11
3	CAL3	1	Water			10.03	04/07/2017	14:12
4	ICV	1	Water			6.99	04/07/2017	14:13
5	CCV1	1	Water			2.02	04/07/2017	14:15
6	I2595-08	1	Water			10.22	04/07/2017	14:16
7	I2595-08DUP	1	Water			10.23	04/07/2017	14:17
8	CCV2	1	Water			12.01	04/07/2017	14:18

LB 86760

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I2595 WorkList ID : 97246 Date : 4/7/2017 1:53:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/07/2017	Water	I2595-08	Corrosivity		PSEG01		A2394	04/07/2017	9040C

Date/Time 04/07/17 11:53 AM
Received by: JA
Relinquished by: CR

Date/Time 04/07/17 3:30 PM
Received by: CP
Relinquished by: JA