



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB86762

Temp.: 20.5 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	04/07/2017	14:10
2	CAL2	1	Water			7.00	04/07/2017	14:11
3	CAL3	1	Water			10.03	04/07/2017	14:12
4	ICV	1	Water			7.00	04/07/2017	14:13
5	CCV1	1	Water			2.02	04/07/2017	14:22
6	I2595-03	1	Solid	20.03	20	8.58	04/07/2017	14:23
7	I2595-06	1	Solid	20.02	20	8.33	04/07/2017	14:25
8	I2595-06DUP	1	Solid	20.04	20	8.34	04/07/2017	14:26
9	CCV2	1	Water			12.01	04/07/2017	14:27

WORKLIST(Hardcopy Internal Chain)

1386762

WorkList Name : CORROSIVITYS I2595

WorkList ID : 97249

Date : 4/7/2017 1:54:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/14/2017	Solid	I2595-03	Corrosivity		PSEG01		GRID-1	04/07/2017	9045C
04/14/2017	Solid	I2595-06	Corrosivity		PSEG01		GRID-4	04/07/2017	9045C

Date/Time 4/10/17 15:03
 Received by: JD
 Relinquished by: JD

Date/Time 04/07/17 3:30 PM
 Received by: JD
 Relinquished by: JD